

AGL ^{GRUP}

at your service globally.

NitrAG[®]



DISPOSABLE POWDER FREE
EXAMINATION **GLOVE**

PERSONAL PROTECTIVE EQUIPMENT
A PROTECTION AND SAFETY SUPPLIER

www.aglgrup.com.tr

www.nitrage.com.tr



SOFT 
TOUCH

RANGE OF APPLICATION



▲ RESTAURANTS



▲ HOSPITALS



▲ HEALTHCARE INSTITUTIONS



▲ SHOPPING MARKETS



▲ HOME TASKS



▲ URBAN LIFE



▲ OFFICES



▲ CABINE CREWS



▲ SCHOOLS

FEATURES

The Soft Touch Termaplus Poly gloves prevent any allergic reactions among workers and customers. Made from polyethylene (LDPE) with 7.5/8 Micron, the gloves do not release any unwanted chemicals and therefore allow direct contact with food or any other sensitive items.

As an excellent alternative to expensive gloves made from various other materials, NitrAG gloves are suitable for safety use in Food & Beverage industries, Consumer Services as SPA, beauty centers, barbers and daily life.



- ✓ Excellent fit, maximum sensitivity and dexterity for fine tasks. The entire glove is finely textured (including palms and fingers) to provide consistent grip on glassware, small objects, instruments and tools.
- ✓ Isolate odors and dust to keep your hands clean
- ✓ Ideal for all kinds of uses that require a safe and hygienic environment
- ✓ Ideal for almost any application including tattooing, food preparation, painting, cleaning, pet care, home improvement, hobbies, arts and crafts
- ✓ Convenient dispense pack keeps gloves clean and organized while providing easy access

DISPOSABLE POWDER FREE **GLOVE**


NitrAG[®]

PACKAGING FORM

AVAILABLE COLOURS



AVAILABLE SIZES



100 GLOVES



24 BOXES



72 CARTONS



33 PALLETS

1 Box 100 PCS | 1 Carton 24 Box | 1 Palatte 72 Carton | 1 Truck 57.024 Box

SOFT TOUCH



TECHNICAL DATA SHEET

Artical Number	130001 (S), 130002 (M), 130003 (L), 130004 (XL)
Packaging	100 PCS/Box 24 Boxes/Carton 2400 PCS/Carton
Glove Specifications	Material: TPE Color: Blue Surface: Texturierte Fingerstip Cuff Design: Rolled Edge Latex: Free
Thickness	Fingertips: 0,7 mm
Physical Properties	EN 455-1 Freedom From Holes: 1,5 AQL EN 455-2 Physical Properties : Long, 240 mm Width, 110 mm Dimensions : Force and break: 6,5 N Force and break after : 6,2 N EN 455-3 Powder Fee : 0,3 mg Per Glove
Chemical Resistance for Degradation	TS EN 374-4 + EN 374-1
Resistance to Permation by Chemicals	EN 374-4, TS EN 16523-1, TS EN 659
PPE Certification	Module B Catagory 3
Virus Penetration	BS ISO 16604 + TS EN 14126, Part4.1.4.1 + EN 374-5, Part 5.3
Color and Sizes	BLUE SMALL MEDIUM LARGE X-LARGE
Dimensions	Box: B 196 mm x 110 mm x 48 mm Carton: B 420 mm x 235 mm x 310 mm
Directive	Prepared according to 2016/425 Personal Protective Equipment Directive
Expiration Time	36 Months after production
Country of Manufacture	TURKEY

DISPOSABLE POWDER FREE **GLOVE**

NitrAG[®]



MEDIUM **BLUE**



LARGE **BLUE**



X-LARGE **BLUE**



SMALL **BLUE**

SOFT TOUCH

DISPOSABLE POWDER FREE **GLOVE**

NitrAG[®]



MEDIUM **BLACK**



LARGE **BLACK**



X-LARGE **BLACK**



SMALL **BLACK**

SOFT TOUCH

DISPOSABLE POWDER FREE **GLOVE**

NitrAG[®]



MEDIUM **WHITE**



LARGE **WHITE**



X-LARGE **WHITE**



SMALL **WHITE**

SOFT TOUCH

DISPOSABLE POWDER FREE **GLOVE**

NitrAG[®]



MEDIUM **PINK**



LARGE **PINK**



X-LARGE **PINK**



SMALL **PINK**

SOFT TOUCH

EU DECLARATION OF CONFORMITY

AGL GRUP

Declaration Number : 208-21-01-R02-01

The manufacturer : AGL GRUP PETROL KOZMETİK MEDİKAL İNŞ. GIDA SAN. TİC. LTD. ŞTİ.
TUGAY YOLU CD. OFİSİM İSTANBUL PLAZA B BLOK NO: 88 KAT:15
34844 MALTEPE, İSTANBUL, TURKEY

Declare that the product : TPE disposable powder free examination and protective gloves **NitrAG PLUS, NitrAG**

Directive : Prepared according to 2016/425 Personal Protective Equipment Directive

Class : 3

Module : B

These are EN4551-2-3 and EN 374 1-2-3-4-5 ,CE 2841 class 3 approved gloves and are an alternative to nitrile. Manufactured with the new generation technology and sets itself apart from any other product in the industry. The silver ions it contains allow for not only better touch sensitivity but also a more comfortable prolonged use. Being powder-free attracts demand for consumers looking to use anti-allergic products in various domains including healthcare, food, home and sanitary.

Our Thermaplus Poly gloves prevent any allergic reactions amongst users and customers. Made from polyethylene (LDPE) with 7.5/8 Micron, the gloves do not release any unwanted chemicals and therefore allow direct contact with food or any other sensitive items.

Our gloves are suitable for use safely in domestic and commercial environments including Food & Beverage industries, Healthcare & Hospitality, Spas, Beauty Centres, Hairdressers, Barbers, etc.

Benefits of the gloves:

- ✓ Excellent fit, maximum sensitivity and dexterity for fine tasks. The entire glove is finely textured (including palms and fingers) to provide consistent grip on glassware, small objects, instruments and tools.
- ✓ Isolate odours and dust to keep your hands clean
- ✓ Ideal for all kinds of uses that require a safe and hygienic environment
- ✓ Ideal for almost any application including tattooing, food preparation, painting, cleaning, pet care, home improvement, hobbies, arts and crafts
- ✓ Convenient dispense pack keeps gloves clean and organized while providing easy access

Colours available: White, Black, Blue & Pink
Sizes: S-M-L-XL

100 gloves in a box

Is conformal to the following directives and standards:

- EN 420 Protective Gloves
- EN ISO 374-1 Protective Gloves against hazardous chemicals and microorganisms
- EN ISO 374-5 Protective Gloves against dangerous chemicals and micro-organisms - Part 5: Terminology and performance requirements for micro-organisms risks
- EN 455-2 Medical gloves for single use - Part 2: Requirements and testing for physical properties
- EN 455-3 Medical gloves for single use - Part 3: Requirements and testing for biological evaluation

CE 2841

AGL GRUP PETROL KOZMETİK
MEDİKAL İNŞ. GIDA SAN. TİC. LTD. ŞTİ.
Cumhuriyetiç. Söğütözü Sk. H. Sebep Apt.
No: 88 Kat: 15 / İSTANBUL
TARİHİ VERGİ Dairesi: 009 144 6578

SOFT TOUCH

NitrAG® NitrAG® NitrAG®



AB Tip İnceleme Sertifikası EU Type-Examination Certificate

Belge No / Certificate No : 208-21-01-R02
 Belgeleme Tarihi / Belge Sonuğu Tarihi /
 Certification Date / Certificate Validity Date : 22.04.2021-31.03.2026
 Belge Geçerlilik Tarihi / Document Validity Period : 5 yıl / 5 years
 Firma Unvanı ve Adresi /
 Company Name and Address :

AGL GRUP PETROL, KOZMETİK, MEDİKAL, İNŞAAT
 GIDA SAN. TIC. LTD. ŞTİ.
 Tugay Yolu Cad. Ofisim İstanbul Plaza B Blok No: 88 Kat:
 15 Mahallesi İSTANBUL.

Ürün Adı /Modeller / Product Name / Models :

NitrAGPlus
 NitrAG

Direktif / Directive
 Modül /Kategori / Module / Category :

2016/425 REGULATION
 B MODÜLÜ / KATEGORİ III
 MODULE B - CATEGORY III
 MNA-M-2021-00293

Test Rapor No / Test Report No
 Ürün Tipi / Product Type :

- EN 420+A1 Korumaya Eldivenler / Protective gloves
- EN ISO 374-1 Tehlikeli Kimyasallar Ve Mikroorganizmalara Karşı Korumaya Eldivenler (Performans Seviyeleri: Tip C) / Protective Gloves Against Dangerous Chemicals And Micro-Organisms (Performance Level: Type C)
- EN ISO 374-5 Tehlikeli Kimyasallar Ve Mikroorganizmalara Karşı Korumaya Eldivenler - Bölüm 5: Mikroorganizmal riskler için terim ve performans kriterleri / Protective Gloves Against Dangerous Chemicals And Micro-Organisms Part 5: Terminology and performance requirements for micro-organisms risks

Ürünün Malzeme Bilgisi / Product Material Information: NitrAGPlus, NitrAG model ürünleri termoplastik elastomer kullanılarak üretilmiştir. / NitrAGPlus, NitrAG model products are manufactured using thermoplastic elastomer.

Revizyon nedeni / Reason for revision: Model adı ve firma adresi revize edilmiştir. / The model name and company address have been revised.

Volkan AKIN
 22.04.2021
 Karar Verici / Approver

Okan AKEL
 22.04.2021
 Şirket Müdürü / General Manager



MNA Laboratuvarları San. Tic. Ltd. Şti.
 Adres: 0216 Mahallesi Mihalim Yarıdağlar Cad. No:21 Akapazar / İstanbul
 Tel: 0216 574 07 08 Faks: 0216 575 13 13 www.mnalab.com

U-Form-002/Rev.04/17.09.2020

NitrAG® NitrAG® NitrAG®



Notified Body Number: 2841

ATTACHMENTS (208-21-01-R02)

To certify the PPE product at Category III level, C2 or D module is accompanied by applying one of the conformity assessment methods along with the EU Type Examination (Module B).

Model : NitrAGPlus, NitrAG

PPE SPECIFICATION	PERFORMANCE LEVELS
Dexterity	5
Material Resistance To Permeation By Chemicals	1 (Type C)
Phi-X174 Bacteriophage	Appropriate
Degradation (EN ISO 374-4:2019)	%40 NaOH -5,65 %

PPE produced as a single unit to fit an individual user, all the necessary instructions for manufacturing such PPE on the basis of the approved basic model:

MARKING		
MANUFACTURER: AGL GRUP PETROL, KOZMETİK, MEDİKAL, İNŞAAT GIDA SAN, TIC. LTD. ŞTİ.		
PPE TYPE:		
<ul style="list-style-type: none"> - EN 420+ A1 Protective gloves - EN ISO 374-1 Protective Gloves Against Dangerous Chemicals And Micro-Organisms - EN ISO 374-5 Protective Gloves Against Dangerous Chemicals And Micro-Organisms Part 5: Terminology and performance requirements for micro-organisms risks 		
PRODUCT SIZE / MODEL: NitrAGPlus, NitrAG (S, M, L, XL)		
PICTOGRAM AND PERFORMANCE LEVELS:		
EN 420+A1	EN ISO 374-1/2016 Type C	EN ISO 374-5/2016
NB 2841		

MNA LABORATUVARLARI SAN, TIC. LTD. ŞTİ declares that the above-mentioned product meets the requirements of the directive according to the EU Directive 2016/425, the safety of the product is covered by the conditions and use specified in this certificate and in the technical file.

MNA Laboratuvarları San. Tic.Ltd. Şti
Adres: Küçükbakkalköy Mahallesi Yenidoğan Cad.No:21 Atajehir/İstanbul
Tel: 0216 524 07 08 Faks: 0216 525 13 31 www.mnalab.com

U.Firm:007/Rev:04/12.03.2020

Page 1 / 2


Notified Body Number: 2841

ATTACHMENTS (208-21-01-002)



DOCUMENTS IN THE TECHNICAL FILE


- Basic Health Safety Requirements
- Risk Assessment
- Test Reports
- Technical Report

MNA Laboratuvarları San. Tic. Ltd. Şti.
Adres: Kışıköy Mahallesi Yemirgen Cad.No:21 Ataşehir/İstanbul
Tel: 0236 574 07 08 Faks: 0236 575 13 35 www.mnalab.com


9-Farm-001/Rev.04/12.01.2020

Page 2 / 2

NitrAG® NitrAG® NitrAG®

	MNA LABORATUVARLARI
TECHNICAL EVALUATION REPORT (208-21-01-R02)	

Report No :208-21-01-R02
Report Date :22.04.2021
Application No :208-21-01

- 1. COMPANY INFORMATION:**
AGL GRUP PETROL KOZMETIK MEDİKAL HİŞAAT GIDA SAN. TIC. LTD. ŞTİ.
Teşay Yolu Cad. Ofisim İstanbul Plaza B Blok No: 88 Kat: 15 Maltepe/ İSTANBUL
Tel: +90 216 457 13 00
Mail: info@aglgrup.com.tr
- 2. PPE INFORMATION:**
Disposable and non-sterile thermoplastic elastomer glove.
- 3. PPE TYPE IDENTIFICATION**
EN 420+A1 Protective gloves
EN ISO 374-1 Protective Gloves Against Dangerous Chemicals And Micro-Organisms
EN ISO 374-5 Protective Gloves Against Dangerous Chemicals And Micro-Organisms Part 5: Terminology and performance requirements for micro-organisms risks
- 4. PPE PICTURES**

NitrAG Plus
NitrAG
- 5. PPE DIMENSIONS:**
NitrAG Plus, NitrAG model product has been found to be produced using S, M, L, XL sizes.
- 6. PPE PRODUCT MATERIAL INFORMATION:**
The product is made of thermoplastic elastomer.

© FEM-006 REV.00 FAYDASI 08.11.2019 Page 1 / 3

NitrAG® NitrAG® NitrAG®



NitrAG® NitrAG® NitrAG®

 <small>MNA LABORATUVARLARI</small> <small>www.laboratuvarlar.com.tr</small>	MNA LABORATUVARLARI	
	TECHNICAL EVALUATION REPORT (208-21-01-R02)	

7. ESSENTIAL HEALTH AND SAFETY REQUIREMENTS

- Visual examination has been made for ergonomics according to EN 420.
- Protection levels and degrees are defined by the manufacturer.
- Suitable construction materials are determined by visual inspection according to EN 420 + A1 Article 4.1.
- pH content is determined according to EN 420 + A1 clause 4.3.2.
- The glove's ability has been tested and evaluated according to EN 420 + A1.
- The measurement of hand and glove has been done, it has been evaluated according to EN 420 + A1.
- The analysis of the glove has been made according to EN 420 + A1 and EN ISO 374-1. Protective gloves against hazardous chemicals and microorganisms have been evaluated according to EN ISO 374-1.
- Determination of Organotin Compounds (DOT In-House Method: modified from EN ISO 16179), Phthalate Determination (ISO / TS 16181) are analyzed according to the standard.
- Analyzes against harmful chemicals have been performed and evaluated according to the REACH regulation.
- Gloves have been analyzed and evaluated according to the requirements of EN ISO 374-5 standard against microbial risks.

**B. ANALYSIS AND EVALUATIONS:
EN 420+A1**

ANALYSIS	PERFORMANCE LEVEL					RESULT	PERFORMANCE LEVEL	EVALUATION
	1	2	3	4	5			
pH	3,5<value<9,5					7,12	3,5<value<9,5	PASS
Dexterity	11 mm	9,5 mm	8 mm	6,5 mm	5 mm	5 mm	5	PASS
Organotin Compounds (DOT)	<1000 ppm					<10 ppm	<1000 ppm	PASS
Phthalates	<1000 ppm					<50 ppm	<1000 ppm	PASS

Size	Circumference (mm)	Length (mm)	RESULT		EVALUATION
			Circumference (mm)	Length (mm)	
9	229	192	301	210	PASS

* The product is produced according to special dimensions.

Size	Glove length (mm)	RESULT	EVALUATION
9	250	255	PASS

* The product is produced according to special dimensions.

NitrAG® NitrAG® NitrAG®



MNA LABORATUVARLARI
TECHNICAL EVALUATION REPORT (208-21-01-R02)

EN ISO 374-1

ANALYSIS	PERFORMANCE LEVEL	RESULT	EVALUATION
Part 4: Determination of resistance to degradation by chemicals	There is no performance value. Only reporting is made.	76-5,65	
Part 2: Determination of resistance to penetration (Air leak test)	No leak to be detected	No leak	PASS
Part 2: Determination of resistance to penetration (Water leak test)	No leak to be detected	No leak	PASS
Determination of material resistance to permeation by chemicals	2 (>30 min no leak. Sodium Hydroxide 40 %)	No leak	PASS

EN ISO 374-5

ANALYSIS	PERFORMANCE LEVEL	RESULT	EVALUATION
Clothing for protection against contact with blood and body fluids.	No leak to be detected according to ISO 16604 Procedure B.	No leak (0 PFL/ml)	PASS

9. DECISION PROPOSAL

Analysis and examinations NitrAGPlus, NitrAG model coded personal protective equipment; EN 420+A1 Protective gloves; EN ISO 374-1 Protective Gloves Against Dangerous Chemicals And Micro-Organisms; EN ISO 374-5 Protective Gloves Against Dangerous Chemicals And Micro-Organisms Part 5; Terminology and performance requirements for micro-organisms risks, Experiments and Marking standards are evaluated. It is recommended to be certified at the performance levels specified as a result of technical evaluations.

10. ATTACHMENTS

- Basic Health Safety Requirements
- Risk Assessment
- User Instruction

Reason for revision : The model name and company address have been revised.

CONTROLLER : VOLKAN AKIN

SIGN : 

DATE : 22.04.2021

Page 3 / 3

mna CONFORMITY TO TYPE BASED ON INTERNAL
PRODUCTION CONTROL PLUS SUPERVISED PRODUCT
CHECK AT RANDOM INTERVALS
(MODULE C2, ANNEX VII) (208-21-01-R02-01)

Notified Body Number: 2841


Report No : 208-21-01-R02-01
Report Date : 22.04.2021
Application No : 208-21-01-R02-01

1. COMPANY INFORMATION:
AGL GRUP PETROL KÖZMETİK MEDİKAL İNŞAAT GIDA SAN. TİC. LTD. ŞTİ
Tugay Yolu Cad. Ofisim İstanbul Plaza B Blok No: 88 Kat: 15 Maltepe/ İSTANBUL
Tel: +90 216 457 13 00
Mail: info@aglgrup.com.tr

2. PPE INFORMATION:
Disposable and non-sterile thermoplastic elastomer glove.

3. PPE TYPE IDENTIFICATION
EN 420+A1 Protective gloves
EN ISO 374-1 Protective Gloves Against Dangerous Chemicals And Micro-Organisms
EN ISO 374-5 Protective Gloves Against Dangerous Chemicals And Micro-Organisms Part 5: Terminology and performance requirements for micro-organisms risks

4. PPE PICTURES



NitrAGPlus
NitrAG

5. PPE DIMENSIONS:
NitrAGPlus, NitrAG model product has been found to be produced using S, M, L, XL sizes.

6. PPE PRODUCT MATERIAL INFORMATION:
The product is made of thermoplastic elastomer.

U.FRA05615/00.YAYIN TARİHİ: 20.11.2023 Page 1 / 3

NitrAG® NitrAG® NitrAG®



Notified Body Number: 2841

CONFORMITY TO TYPE BASED ON INTERNAL
 PRODUCTION CONTROL PLUS SUPERVISED PRODUCT
 CHECK AT RANDOM INTERVALS
 (MODULE C2, ANNEX VII) (208-21-01-R02-01)

7. ESSENTIAL HEALTH AND SAFETY REQUIREMENTS

- Visual examination has been made for ergonomics according to EN 420.
- Protection levels and degrees are defined by the manufacturer.
- Suitable construction materials are determined by visual inspection according to EN 420 + A1 Article 4.1.
- pH content is determined according to EN 420 + A1 clause 4.3.2.
- The glove's ability has been tested and evaluated according to EN 420 + A1.
- The measurement of hand and glove has been done. It has been evaluated according to EN 420 + A1.
- The analysis of the glove has been made according to EN 420 + A1 and EN ISO 374-1. Protective gloves against hazardous chemicals and microorganisms have been evaluated according to EN ISO 374-1.
- Determination of Organotin Compounds (DOT In-house Method: modified from EN ISO 16179), Phthalate Determination (ISO / TS 16181) are analyzed according to the standard.
- Analyses against harmful chemicals have been performed and evaluated according to the REACH regulation.
- Gloves have been analyzed and evaluated according to the requirements of EN ISO 374-5 standard against microbial risks.

8. ANALYSIS AND EVALUATIONS:
EN 420+A1

ANALYSIS	PERFORMANCE LEVEL					RESULT	PERFORMANCE LEVEL	EVALUATION
	1	2	3	4	5			
pH	3.5<value<9.5					7,15	3.5<value<9.5	PASS
Dexterity	11 mm	9,5 mm	8 mm	6,5 mm	5 mm	5 mm	5	PASS
Organotin Compounds (DOT)	<1000 ppm					<10 ppm	<1000 ppm	PASS
Phthalates	<1000 ppm					<50 ppm	<1000 ppm	PASS

Size	Circumference (mm)	Length (mm)	RESULT		EVALUATION
			Circumference (mm)	Length (mm)	
S	229	192	302	211	PASS

* The product is produced according to special dimensions.

Size	Glove length (mm)	RESULT	EVALUATION
S	250	251	PASS

* The product is produced according to special dimensions.

U.FRM.056/REV.00/FORM TARIH:2011.2013

Page 2 / 3

NitrAG® NitrAG® NitrAG®

NitrAG[®] NitrAG[®] NitrAG[®]

AGL[®] GRUP[®]

NitrAG[®] NitrAG[®] NitrAG[®]

**TECHNICAL DOCUMENTATION FOR PPE
TYPE EXAMINATION AND PROTECTIVE
GLOVES ACCORDANCE WITH
STANDARDS EN 420:2003+A1:2009 AND
EN ISO 374-1:2016/A1:2018**

NitrAG[®] NitrAG[®] NitrAG[®]

AGL[®] GRUP[®]

ID: TCF-D- PFN -2020
Version: 01
Issued date: 05.09.2020

NitrAG[®] NitrAG[®] NitrAG[®]

AGL[®] GRUP[®]

NitrAG[®] NitrAG[®] NitrAG[®]

AGL[®] GRUP[®]

AGL GRUP PETROL KOZMETIK
MEDICAL INS. GIDA SAN. TIC. LTD. STI.
Cumhuriyet Mah. Söğütözü Sok. 16. Şişli/Şişli
No: 14/1 Kat:1 / İSTANBUL
Ticaret Sicil No: 272100
Vergi No: 3450000000000000

NitrAG[®] NitrAG[®] NitrAG[®]

NitrAG[®] NitrAG[®] NitrAG[®]
EXAMINATION AND PROTECTIVE GLOVES
ID: TCF-D- PFN -2020
Version: 01
Issued date: 05.09.2020

TABLE OF CONTENT

0. PURPOSE

1. GENERAL REQUIREMENTS

- 1.1 Design general principles
- 1.2 Innocuousness
- 1.3 Comfort and effectiveness
- 1.4 Manufacture's instruction and information

2. ADDITIONAL REQUIREMENTS

- 2.1 Design and fitting systems
- 2.2 Incorporating adjustment systems
- 2.3 Ageing
- 2.4 Protection against cutaneous and ocular contact

3. SCOPE OF APPLICATION OF RECOMENDATION FOR USE EN 420:2003+A1:2009 AND EN ISO 374-1:2016/A1:2018

4. GENERAL REQUIREMENTS OF RECOMENDATION FOR USE EN 420:2003+A1:2009 AND EN ISO 374-1:2016/A1:2018

5. SPECIFIC REQUIREMENTS OF RECOMENDATION FOR USE EN 420:2003+A1:2009 AND EN ISO 374-1:2016/A1:2018

6. SAMPLES SUBMITTED

7. ESSENTIAL REQUIREMENTS

8. RISK ASSESMENT

9. MEANS OF CONTROL

10. MARKING

11. INFORMATION LEAFLET

AGL GRUP PETROL KOZMETIK
MEDICAL INS. GIDA SAN TIC.LTD. STI.
Cumhuriyet Mah. Sultaniye Sk. H. Setbaş Apt.
No: 45/1 Kat:1 / İSTANBUL
Tic Sicil No: 271100 / Vergi No: 009 144 6578

NitrAG® NitrAG® NitrAG®
 EXAMINATION AND PROTECTIVE GLOVES
 ID: TCF-D- PFN -2020
 Version: 01
 Issued date: 05.09.2020

0. PURPOSE

THE PPE TYPE EXAMINATION AND PROTECTIVE GLOVES, designed with the intention of protecting the user against dangerous chemicals, is manufactured by AGL GRUP PETROL KOZMETİK MEDİKAL İNŞAAT GIDA SAN. TIC. LTD. ŞTİ. at the address of Cumhuriyet Mah. Güldünya Sk. H.Şebap Apt. No: 119 Kartal/ İSTANBUL with the general health and safety requirements specified in Regulation (EU) 2016/425, in particular, the specifications contained in standards EN 420:2003+A1:2019 and EN ISO 374-1:2016/A1:2018 published by the European Committee for Standardisation, as CATEGORY III PPE.

1. GENERAL REQUIREMENTS

1.1 Design principles

The PPE are protective gloves which are used to protect the user against dangerous chemicals. The protective gloves are designed and manufactured so that under normal conditions of use without exposure to additional risks, except in the event of the user's individual hypersensitivity, the wearers can perform the activity as normally as possible with appropriate protection.

The ergonomic design takes into consideration the activities that the wearers might perform under normal conditions of use without exposure to additional risks, except in the event of the user's individual hypersensitivity.

1.2 Innocuousness

The gloves' materials and their decomposition products do not adversely affect user hygiene or health. Specifically, the contact surface of the glove with the hand is free of roughness, sharp edges, projection and the likes which may cause excessive irritation or injuries. These gloves have good finger dexterity and do not impede the user movement or sensory perception.

1.3 Comfort and effectiveness

The gloves are light, yet strong under foreseeable condition of use. It offers better puncture resistance than similar natural rubber latex glove.

The pH of the gloves is close to neutral and does not cause irritation.

1.4 Manufacture's instruction and information

See Annex II for the information to be included.

AGL GRUP PETROL KOZMETİK
 MEDİKAL İNŞ. GIDA SAN. TİC. LTD. ŞTİ.
 Cumhuriyet Mah. Güldünya Sk. H.Şebap Apt.
 No: 119 Kartal / İSTANBUL
 YAKARLIK VERGİ ZARFI: 009 144 6579

NitrAG® NitrAG® NitrAG®
 EXAMINATION AND PROTECTIVE GLOVES
 ID: TCF-D- PFN -2020
 Version: 01
 Issued date: 05.09.2020

2. ADDITIONAL REQUIREMENTS

THE PPE TYPE EXAMINATION AND PROTECTIVE GLOVES complies with the design and fitting systems, adjustment systems, ageing and protection against cutaneous and ocular contact defined below.

2.1 Design and fitting systems

The PPE's design and fitting systems enable it to adapt to the morphology of the user as shown in the following documents:

- Description of the PPE in accordance with Annex I.
- Specification of materials and components in accordance with Annex I.

2.2 Ageing

The obsolescence deadline is 3 years from the date of manufacture, the month and year of obsolescence is indelibly and unambiguously marked on its packaging.

2.3 Protection against cutaneous and ocular contact

The gloves intended to prevent the surface contact of the hands with substances and mixtures which are hazardous to health must be capable of preventing the penetration or permeation of such substances and mixtures and agents through the protective integument under the foreseeable conditions of use for which the gloves are intended.

3 SCOPE OF APPLICATION OF RECOMENDATION FOR USE EN 420:2003+A1:2009 AND EN ISO 374-1:2016/A1:2018

A EXAMINATION AND PROTECTIVE GLOVES can only be approved when its individual components satisfy the requirements of the test specifications which may be a complete standard or part of it, and practical behavior tests have been carried out satisfactorily on the complete equipment, as specified in the appropriate standard.

Specifically, these gloves is in accordance with the requirements of EN 420:2003+A1:2009 AND EN ISO 374-1:2016/A1:2018 for protective gloves intended to protect the user against dangerous chemicals and defines terms to be used.

AGL GRUP PETROL KOSMETIK
 MEDIKAL INS. GIDA SAN. TIC. LTD. STI.
 Cumhuriyet An. 5000m2 Sok. H Sebap Apt.
 No: 101 Kat:1 / ISTANBUL
 Yabancı Vergi Dairesi: 009 144 6578

NitrAG® NitrAG® NitrAG®
 4



EXAMINATION AND PROTECTIVE GLOVES
ID: TCF-D- PFN -2020
Version: 01
Issued date: 05.09.2020

4 GENERAL REQUIREMENTS OF RECOMENDATION FOR USE EN 420:2003+A1:2009 AND EN ISO 374-1:2016/A1:2018

4.1 Design:

The protective gloves cover and protect the hands from the wrist to the fingers. These gloves are intended to be used only as a physical barrier against brief contact with chemicals, and they need to be removed and discarded immediately after they become contaminated.

a= Glove Length
b= Hand Length
c= Peripheral Length

1) WERGARD, NitrAG

Available sizes: 6, 7, 8, 9, 10, 11

SIZE (minimum)	6	7	8	9	10	11
GLOVE LENGTH(a)(mm)	220	230	240	250	260	270
HAND LENGTH(b)(mm)	160	171	182	192	204	215
PERIPHERAL LENGTH(c)(mm)	152	178	203	229	254	279

5 SPECIFIC REQUIREMENTS OF RECOMENDATION FOR USE EN 420:2003+A1:2009 AND EN ISO 374-1:2016/A1:2018

5.1 Packaging

Protective gloves are supplied for sale packed in such a way that they are protected against mechanical damage and contamination before use.

AGL GRUP PETROL, KOZMETIK
MEDİKAL İHS. GIDA SAN. TİC. LTD. ŞTİ.
Cumhuriyet Yolu, Gölbaşı Yık. M. Suluca Apt.
No: 145 Kat: 1 / İSTANBUL
Yatırım Yatırım Yatırım: 029 144 6579

NitrAG[®] NitrAG[®] NitrAG[®]
EXAMINATION AND PROTECTIVE GLOVES
ID: TCF-D- PFN -2020
Version: 01
Issued date: 05.09.2020

5.2 Material

The materials used is suitable to withstand handling and use during the period of time for which protective gloves have been designed.

5.3 Finishing of the parts

The parts of the device that will be in contact with the wearer have not sharp edges or burrs.

5.4 Measurement of glove length

Before testing, samples shall be conditioned for at least 24h at $23 \pm 2^{\circ}\text{C}$ and $50 \pm 5\%$ relative humidity and testing shall be started within 10 min after removal from conditioning.

For each size, 3 gloves shall be tested. Measure the length by freely suspending the glove with the middle finger on a vertical graduated ruler having a rounded tip so as to fit the shape of the finger tip of the glove. Remove wrinkles and folds without stretching the glove. Turn the glove around the pin and record the minimum measured length to the nearest millimetre.

5.5 Penetration

Protective gloves shall not leak when tested according to EN 374-2:2014, 7.2 and 7.3.

5.6 Degradation

The degradation shall be determined according to EN 374-4 for each chemical claimed in the marking and reported in the user instruction.

5.7 Permeation

5.7.1 General

All the results should be reported in the user instruction.

Each combination of protective glove/ test chemical shall be classified according to Table 1, using the results as given in EN 16523-1:2015, 8.5.1.1 or 8.5.1.3 for the normalized breakthrough time.

AGL GRUP PETROL KOSMETIK
MEDİKAL İNS. GIDA SAN. TİC. LTD. ŞTİ.
Cumhuriyet Mah. Gölbaşı Sk. H. Sektör Apt.
No: 16/5 Kat:1 / İSTANBUL
YAKARAK VERİLEN ZARFI: 009 144 6578

EXAMINATION AND PROTECTIVE GLOVES
ID: TCF-D- PFN -2020
Version: 01
Issued date: 05.09.2020

Table 1 - Permeation performance levels

Measured breakthrough time (min)	Permeation performance level
>10	1
>30	2
>60	3
>120	4
>240	5
>480	6

The test chemical(s) shall be taken from the list of test chemicals in Table 2. Other test chemicals could be used depending on the application of the gloves.

Table 2- List of test chemicals

CODE LETTER	CHEMICAL	CAS NUMBER	CLASS
A	Methanol	67-56-1	Primary alcohol
B	Acetone	67-64-1	Ketone
C	Acetonitrile	75-05-8	Nitrile compound
D	Dichloromethane	75-09-2	Chlorinated hydrocarbon
E	Carbon disulphide	75-15-0	Sulphur containing organic compound
F	Toluene	108-88-3	Aromatic hydrocarbon
G	Diethylamine	109-89-7	Amine
H	Tetrahydrofuran	109-99-9	Heterocyclic and ether compound
I	Ethyl acetate	141-78-6	Ester
J	n-Heptane	142-82-5	Saturated hydrocarbon
K	Sodium hydroxide 40%	1310-73-2	Inorganic base
L	Sulphuric acid 96%	7664-93-9	Inorganic mineral acid, oxidizing
M	Nitric acid 65%	7697-37-2	Inorganic mineral acid, oxidizing
N	Acetic acid 99%	64-19-7	Organic acid
O	Ammonium hydroxide 25%	1336-21-6	Organic base
P	Hydrogen peroxide 30%	7722-84-1	Peroxide
S	Hydrofluoric acid 40%	7664-39-3	Inorganic mineral acid
T	Formaldehyde 37%	50-00-0	Aldehyde

The situation described in EN 16523-1:2015, 8.5.1.4 is considered a fail due to non-homogeneity of the samples.

According to the permeation performance, chemical protective gloves are classified into three types: type A, type B, type C.

AGL GRUP PETROL KOSMETIK
MEDİKAL İNS. GIDA SAN. TİC. LTD. ŞTİ.
Cumhuriyet Mah. Şişlihanca Sk. H. Setap Apt.
Mas Nişi Katları / İSTANBUL
Ticaret Sicil No: 009 144 6578

EXAMINATION AND PROTECTIVE GLOVES
ID: TCF-D- PFN -2020
Version: 01
Issued date: 05.09.2020

5.7.2 Type A

The permeation performance shall be at least level 2 against a minimum of six test chemical listed in Table 2.

5.7.3 Type B

The permeation performance shall be at least level 2 against minimum of three test chemical listed in Table 2.

5.7.4 Type C

The permeation performance shall be at least level 1 against minimum of one test chemical listed in Table 2.

5.7.5 Requirement for glove type A, B and C

The requirements are mentioned in Table 3.

Table 3 - Requirements for different protection types of gloves

	5.1	5.2	5.4.2	5.4.3	5.4.4
Type A	X	X	X		
Type B	X	X		X	
Type C	X	X			X


X = required

6. SAMPLES SUBMITTED

The following were submitted:

- 50 pairs type EXAMINATION AND PROTECTIVE GLOVES

AGL GRUP PETROL KOSMETIK
MEDİKAL İNS. GIDA SAN. TİC. LTD. ŞTİ.
Cumhuriyet No. 5535/5536 Sok. 14 Şişli/Beşiktaş
No: 14/1 Kat: 1 / İSTANBUL
Yatırım Yatırım Yatırım: 009 144 6578



NitrAG® NitrAG® NitrAG®
 EXAMINATION AND PROTECTIVE GLOVES
 ID: TCF-D- PFN -2020
 Version: 01
 Issued date: 05.09.2020

7.2 Quality control system

7.2.1 Incoming Material Inspection

All Incoming raw materials (that will affect the product's quality) that are subjected to inspection are placed under quarantine area in the store.

Incoming materials are selected randomly as per ISO 2859 simpling plan and inspected by Incoming Material Inspector.

Incoming materials that meet In-house specifications are attached with "Passed" Sticker.

Incoming materials that do not meet In-house specifications are quarantined pending management disposition decision.

- The Incoming materials are needed to be screened are attached with "Rejected Sticker and with "SCREEN" stamp on it.
- The Incoming materials that can be "Used as it is (substandard material but without functional implication and actual affect on quality) are attached with "Rejected" Sticker and with "UAI" stamp on it.
- Incoming materials that cannot be used and needed to be returned back to supplier are attached with "Reject" sticker.

All Reject Materials are return immediately to Supplier or stored in "Reject Area" pending return to supplier.

7.2.2 Milling and Compounding of Raw Material

Only the materials that with "Passed" Sticker can be used for milling process. The milled chemical is analyzed by laboratory personnel.

Only the milled chemical that has passed laboratory analysis will be released to be used in Compounding Process.

The milled chemical that has failed laboratory analysis shall be re-milled and quarantined to be used until it passes laboratory analysis.

Only the approved milled chemicals and chemicals with "Passed" sticker are used for compounding process.

The TPE compounds are analyzed by laboratory personnel.

The TPE compounds that pass all the In-house specification will be release to be use in production.

The TPE compounds that fail the In-house specification will be quarantined be used and shall wait for management disposition decision.

AGL GRUP PETROL KOZMETIK
 MEDICAL INS. GIDA SAN. TIC. LTD. STI.
 Cumhuriyet Mh. Çekirge Sok. 14. Sektör A Blok
 No: 10 Kat: 10 / İSTANBUL / 34099
 Yabancı Vergi Davesi: 009 144 6578

EXAMINATION AND PROTECTIVE GLOVES
ID: TCF-D- PFN -2020
Version: 01
Issued date: 05.09.2020

All adjustments made on the TPE compounds will be documented in a work sheet and shall be analyzed by laboratory personnel again.

7.2.3 Production On-line Goods

Each bin of gloves is issued with a Travel Card and is controlled to 8 kgs.

Samples are taken and inspected as per ISO 2859 table I and II-A.

Every 6 bins, grouped as one Lot shall be tested as follow;

Inspect On	Inspection Level	AQL
i) Pin-hole	80 pcs	4.0
ii) Dimension	S2	1.5
iii) Visual Major	80 pcs	4.0
iv) Visual Minor	80 pcs	6.5

The bins that have passed the above inspection will be marked as grade AQL 4.0 on the Travel card.

The bins that had failed the above inspection will be segregated and quarantined and stored in a specified area in store.

The passed bins will proceed to WIP Store or Packing Department.

7.2.4 Packing QC Inspection

At packing, Packing QC conduct line clearance inspection before commence packing and at every hourly.

7.2.5 Production On-line Goods

The pallets of finish packed gloves are inspected by pre-shipment QA.

The samples are randomly picked and inspected as per ISO 2859 table I and II-A for:

Inspect On	Inspection Level	AQL
i) Pin-hole (Watertight)	S4	4.0
ii) Dimension	S2	4.0
iii) Visual Major	S4	4.0
iv) Visual Minor	S4	6.5
v) pH of Glove	N=3	-

The pallets of packed gloves that have passed Pre-shipment QA Inspection ar ready to deliver and store in outgoing store.

The pallets of packed gloves that have failed Pre-shipment QA Inspection shall be reworked by Packing Department.

AGL GRUP PETROL KOSMETIK
MEDIKAL PAS. GIDA SAN TIC. LTD. STI.
Cankaya/Ankara, Cankaya Sk. H. Setip Apt.
No: 46/1 Katilari / ISTANBUL
YAKARCI YEREL TELEFON: 009 344 6379 11

NitrAG® NitrAG® NitrAG®
 EXAMINATION AND PROTECTIVE GLOVES
 ID: TCF-D- PFN -2020
 Version: 01
 Issued date: 05.09.2020

The sorted gloves will be quarantined and stored in specified area and wait management disposition decision.

7.3 Control And Test Facilities

Description of the Control and Test Facilities

Area	Control and Test Facilities
a) Laboratory	1) Chemical concentration testing apparatus.
b) Quality Control (QC) Section	1) Water-tight Test Equipment. 2) Stainless Steel Ruler. 3) Thickness Gauge.
c) Pre-Packing QA Section	1) Water-tight Test Equipment. 2) Stainless Steel Ruler.
d) Pre-shipment QA Section	1) Water-tight Test Equipment. 2) Stainless Steel Ruler. 3) Thickness Gauge.

8 ESSENTIAL REQUIREMENTS

Regulation 2016/425, Annex II	Clause(s)/ sub-clause(s) of EN 420:2003+A1:2009	Clause(s)/ sub-clause(s) of EN ISO 374-1:2016
1.2.1.1 Innocuousness of PPE- suitable constituent materials	4.2	
1.2.1.3 Maximum permissible user impediment	5.2	
1.4 Information supplied by the manufacturer	7.3	7
2.4 PPE subject to ageing	4.3 7.2.1.1 f) & 7.2.2 g)	
2.12 PPE bearing one or more identification or recognition marks directly or indirectly relating to health and safety	7.2.1.1 d) – 7.2.2 e) – 7.3.5	6
3.10 Protection against dangerous substances and infective agents 3.10.2 Protection against cutaneous and ocular contact		5.2, 5.3, 5.4, 5.5

AGL GRUP PETROL KOZMETİK
 MEDİKAL İNS. GIDA SAN. TİC. LTD. ŞTİ.
 Cumhuriyet Mah. Çarşıya Sk. H. Sebep Apt.
 No: 15/1 Kat:1 / İSTANBUL
 YAKARCA Vergi Daresi: 009 144 6578

EXAMINATION AND PROTECTIVE GLOVES
ID: TCF-D- PFN -2020
Version: 01
Issued date: 05.09.2020

9 RISK ASSESSMENT

Risk assessment according to EN ISO 374-1:2016 Prevention of occupational hazards.

RISK AGAINST WHAT IT PROTECTS	TYPE OF WORK	PROBABILITY OF A DANGER / INJURY HAPPENING	QUALIFY THE SERIOUSNESS OF THE DANGER/INJURY	GENERAL CALIFICATION	PROTECTION DEGREE	STANDARD FOR EVALUATING THE LEVEL OF PROTECTION	SOURCE THAT CONTRIBUTES TO THE EXHIBITION
Determination of material resistance to permeation by chemicals	Sanitary ambient, hospitals - laboratory	Very high	High	IMPORTANT RISK	Determination of material resistance to permeation by 1% NaOH (> 30min, PERFORMANCE LEVEL:2	5.2, 5.3, 5.4, 5.5 EN 374-1 7 EN 374-1 7.3 EN 420:2003+A1:2009 6 EN 374-1	Inadequate risk assessment Inappropriate use as indicated in the information leaflet Inappropriate use as indicated in the information glove mark

10 MARKING

Each glove should be marked so that the following information is easily readable by the user and remains readable throughout its intended use:

- 1- Standard EN 374-1/Type C
- 2- Manufacturer's name or brand: AGL GRUP PETROL KOZMETİK MEDİKAL İNŞAAT GIDA SAN. TİC. LTD. ŞTİ.
- 3- PPE reference: WERGARD, NitrAG
- 4- CE marking: CE
- 5- Production control body: 2841
- 6- Classification: CATEGORY III

11 INFORMATION LEAFLET

The leaflet that accompanies each EXAMINATION AND PROTECTIVE GLOVES, an example of which is attached in Annex II, is written in the official language of the Member State in which it is sold, as well as other possible languages.

AGL GRUP PETROL KOZMETİK
MEDİKAL İNŞAAT GIDA SAN. TİC. LTD. ŞTİ.
Cumhuriyetk. Söğütözü Sk. 14. Şişli/İSTANBUL
MKT: 0212 444 4578
Yakarık Veriş Dairesi: 0212 144 4578

NitrAG[®] NitrAG[®] NitrAG[®]
EXAMINATION AND PROTECTIVE GLOVES
ID: TCF-D- PFN -2020
Version: 01
Issued date: 05.09.2020

ANNEX I**PPE DESCRIPTION****1.1 General Description****1.1.1 Glove Type**

EXAMINATION AND PROTECTIVE GLOVES.

Variants : Applicable for smooth or textured or finger-textured surface and meet the product specification.

Standard EN 374-1/Type C

1.1.2 Feature

Ambidextrous, Disposable, Beaded Cuff

1.1.3 Material

Made from Thermoplastic elastomer

1.1.4 Colour

Blue colored

1.1.5 Shelf-Life

3 years from date of manufacture.

1.2 Intended Use

This disposable Personal Protective Equipment (PPE) TPE glove is intended to be worn by an individual for prevention the surface contact of the hands with substances and mixtures which are hazardous to health

1.3 Classification

CATEGORY III

1.4 Applicable Harmonised Standards

- EN 420:2003+A1:2009 Protective gloves – General requirements and test methods
- EN ISO 374-1:2016/A1:2018 - Protective gloves against dangerous chemicals and micro-organisms - Part 1: Terminology and performance requirements for chemical risks

AGL GRUP PETROL KOZMETIK
MEDICAL INS. GIDA SAN. TIC. LTD. STI.
Cumhuriyet Mh. 55. Cad. No: 14, Sultaniye Apt.
No: 14/1 Kat: 1 / ISTANBUL
Tic. Sic. No: 271100 / Vergi Dairesi: 009 144 6579

EXAMINATION AND PROTECTIVE GLOVES
ID: TCF-D- PFN -2020
Version: 01
Issued date: 05.09.2020

1.5 Product Specification

No	Tests	Limit Value	Reference Test method	Evaluation
1	Determination of material resistance to permeation by chemicals (METHANOL)	>10 min	TS EN 16523-1	PERFORMANCE LEVEL 0
2	Dexterity	5 mm	TS EN 420 + A1 Part 6.2	PERFORMANCE LEVEL 5
3	Determination of organotin compounds (DOT)	<1000 ppm	ISO TS 16179	PASS
4	Determination of Phthalates	<1000 ppm	ISO/TS 16181	PASS
5	Determination of pH	3.5 < Result < 9.5	EN 420 + A1 Part 4.3.2 TS EN ISO 3071	PASS
6	Air leak test	Should be no	TS EN 374-2 Part 5.2	PASS
7	Water leak test	Should be no	TS EN 374-2 Part 5.3	PASS
8	Determination of resistance to degradation by chemicals (ACETONE)		TS EN 374-4	
9	Determination of material resistance to permeation by chemicals (%40 NaOH)	> 30 min	TS EN 16523-1	PERFORMANCE LEVEL 2
10	Determination of material resistance to permeation by chemicals (%96 H2SO4)	> 10 min	TS EN 16523-1	PERFORMANCE LEVEL 0
11	Chemical determination of formaldehyde content	75 mg/kg	TS EN ISO 17226-1	PASS

AGL GRUP PETROL KOZMETIK
MEDİKAL İNS. GIDA SAN. TİC. LTD. ŞTİ.
Cemalpaşa Mah. Gökçeçayırı Sk. H. Sebep Apt.
No: 100 Kat:1 / İSTANBUL
Yanık Yeri Dairesi: 009 144 6578

NitrAG[®] NitrAG[®] NitrAG[®]
EXAMINATION AND PROTECTIVE GLOVES
ID: TCF-D- PFN -2020
Version: 01
Issued date: 05.09.2020

ANNEX III: GLOVES MARKING

AGL GRUP
AGL GRUP PETROL KOZMETİK MEDİKAL İNŞAAT
GIDA SAN. TİC. LTD. ŞTİ.
EXAMINATION & PROTECTIVE GLOVES[®]
EN ISO 374-1:2016/TYPE C

CE 2841






AGL GRUP PETROL KOZMETİK
MEDİKAL İNŞ. GIDA SAN. TİC. LTD. ŞTİ.
Cumhuriyet An. 555. Sıra No. 14 Sebepi Kat.
Nispetiye / Beşiktaş / İSTANBUL
Yatırım Vergisi Dairesi: 009 144 6378



EXAMINATION AND PROTECTIVE GLOVES
ID: TCF-D- PFN -2020
Version: 01
Issued date: 05.09.2020

ANNEX IV: PACKAGING

PICTOGRAMS EXPLANATIONS:

-  The permeation performance against Sodium hydroxide 40%
-  yyyy/mm
(usable life)
-  Maximum relativity humidity of storing
-  40°C
10°C
Temperature range of storing
-  See information supplied by the manufacturer

AGL GRUP PETROL,KOZMETIK
MEDİKAL İNS. GIDA SAN. TİC.LTD.ŞTİ.
Cumhuriyet Yolu, Çarşıbaşı Sok. 14, Şilekap. Ağız.
No: 14/11 Kat:1A / İSTANBUL
Yatırım ve Yatırımcı Dairesi: 009 144 6578

19

MNA LABORATORIES
TEST REPORT

Report No:	Date:	Page:	Rev:
M-2021-00293	25.02.2021	1 / 3	
Purpose of Analysis	: SPECIAL REQUEST	Brand	: NitrAG
Sample Type	: GLOVE	Model	:
Sample Send Org.	: AGL GRUP PETROL KOZMETIK MEDİKAL	Sampler	: CUSTOMER
Manufacturer Name	: AGL GRUP PETROL KOZMETIK MEDİKAL		
Analysis Date	: 15.02.2021		
Sample Quantity	: 100 pieces		
Other Informations	:		

No	Tests	Results	Limit Value	Method	Evaluation	Physical Condition
1	Finger Dexterity *	5 (mm)	5 mm	TS EN 420 + A1 Part 6.2	PERFORMANC E LEVEL 5	
2		5 (mm)	5 mm	TS EN 420 + A1 Part 6.2		
3	Determination of Organotin Compounds (DOT) *	<10 (mg/kg)	<1000 ppm	In House Method SOP 05 Rev01 (Modified from ISO TS 16170)	PASS	
4	Determination of Phthalates	<50 (mg/kg)	<1000 ppm	ISO/TS 16183	PASS	
5	Penetration By Blood Borne Pathogen(Bacteriophage)	0 (PFU/ml)	<1 PFU/ml	BS ISO 15604+ TS EN 14326 Part 4.3.4.1 + EN 374-5 Part 5.3	PASS	
6	Determination of pH Textile*	7,32	3.5 + Result + 9.5	EN 420 + A1 Part 4.3.2 TS EN ISO 3071	PASS	
7	Air Leak Test *	No leak	No leak to be	TS EN 374-2 Part 3.2-374-3 Part 4.2	PASS	
8	Water Leak Test *	No leak	No leak to be	TS EN 374-2 Part 3.3-374-3 Part 4.2	PASS	
9	Resistance To Degradation By Chemicals *	-0,65 (NaOH) (%)		TS EN 374-4 + EN 374-1 Part 4.3		
10	Resistance To Permeation By Chemicals *	1,40 NaOH 30 min no leak. (ug/cm ² .min)	> 30 min	TS EN 10523-1 + EN 374-3 Part 4.1	PERFORMANC E LEVEL 2	
11	Sizing and measurement of gloves*	255 (mm)	≥ 250 mm	TS EN 420 + A1 Part 6.1	PASS	
		Hand length 210 (mm)		TS EN 420 + A1 Part 6.1		
		Circumference 301 (mm)		TS EN 420 + A1 Part 6.1		

MNA 23/rev.02/09.08.2019

Küçükçekircekli Mahallesi Yeniböğaziçi Cad. No: 21 Atasehil Bölgesi
info@mna.com.tr tel:02165740308 fax:02165751331

MNA LABORATORIES
TEST REPORT

Report No:	Date:	Page:	Rev:
M-2021-98293	24.02.2021	2 / 3	
Purpose of Analysis	: SPECIAL REQUEST	Brand	: NitrAG
Sample Type	: PROTECTIVE GLOVE	Model	:
Sample Send Org	: AGL GRUP PETROL KOZMETIK MEDİKAL	Sampler	: CUSTOMER
Manufacturer Name	: AGL GRUP PETROL KOZMETIK MEDİKAL		
Analysis Date	: 15.02.2021		
Sample Quantity	: 100 pieces		
Other Informations	:		

	228mm	240mm	TS EN 420 + A3 Part 6.1	FAB
12 Sizing and measurement of gloves*	Hand length	221mm	TS EN 420 + A3 Part 6.1	
	Circumference	205mm	TS EN 420 + A3 Part 6.1	

SAMPLE PLACE

- 1 Line Sample place for finger assembly, Size 9
 12 Line Sample place for sizing and measurement of gloves, Size 10
 11 Line Sample place for sizing and measurement of gloves, Size 9
 2 Line Sample place for finger assembly, Size 10

TMM 25/11/2020 06:42:19

Regulatory Market Supervisor Contact No: 030 445 61 50
info@nitrags.com.tr 0312 670 00 00 / 0312 670 01 11




AB-1163-T
M-2021-03293
26.02.2021

**MNA LABORATORIES
TEST REPORT**

Report No: M-2021-03293	Date: 26.02.2021	Page: 3 / 3	Rev:
Purpose of Analysis : SPECIAL REQUEST	Brand : NitrAG		
Sample Type : PROTECTIVE GLOVE	Model :		
Sample Send Org : AGL GRUP PETROL KOZMETIK MEDİKAL	Sampler : CUSTOMER		
Manufacturer Name : AGL GRUP PETROL KOZMETIK MEDİKAL			
Analysis Date : 15.02.2021			
Sample Quantity : 100 pieces			
Other Informations :			

Operating as an experimental laboratory, MNA Laboratories have been accredited by TÜRKAK with AB-1163-T and TS_EN_950 / IEC_17025_2017 standards. Turkish Accreditation Agency (TÜRKAK) signed a multilateral agreement with the European Accreditation Association (EA) on the recognition of test reports, and a mutual recognition agreement with the International Laboratory Accreditation Association (ILAC). * Analysis is under accreditation.

Note:

- No part of this analysis report can be used alone or separately, and may not be partially copied or reproduced, used to third parties and as a means of advertising without the written permission of the laboratory.
- Analysis results are valid for the above mentioned sample sent by MNA Laboratory company / institution / person. It may not represent the whole.
- Unsigned and unsealed reports are invalid.
- This analysis report cannot be used in judicial-administrative procedures and for advertising purposes.
- Results are valid for the sample as received.
- The decision rule is the rule that determines how measurement uncertainty is taken into account when specifying the PASS density to a specified specification. According to the TLM-052 Decision Rule Implementation instruction, the Decision Rule Implementation Method selected in agreement with CUSTOMER is clearly stated in the report.
- Limit Values are determined by taking from analysis methods.
- The laboratory is not responsible if the information provided by the CUSTOMER affects the validity of the results.
- Test and / or measurement results, expanded measurement uncertainty (if any) and test methods are given in the following pages, which are the supplementary part of this certificate.
- Water Repellency Determination Hydrostatic Pressure Determination TS ISO 811 (Hydrostatic Pressure Tester E / N: 52) Analysis, Seam Strength EN ISO 13968-2 (Strength Test Device E / N: 36) Analysis and resistance to liquid chemical permeation TS EN 858 A1 Part 3.16 (Liquid Chemical Transfer Device E / N: 107) Analysis is carried out in the conditioning room and ISO 159 PART 3.2 conditions (23 ± 2 ° C temperature and 50 ± 4% relative humidity) are applied for ambient conditions.
- List of phthalates analyzed is below.
Di-iso-nonyl phthalate (DINP), CAS number: 28553-12-0 or 68515-48-0
Di-(2-ethylhexyl) phthalate (DEHP), CAS number: 117-81-7
Di-n-octyl phthalate (DNOP), CAS number: 117-84-0
Di-iso-decyl phthalate (DIDP), CAS number: 26761-40-0 or 68515-49-1
Butyl benzyl phthalate (BBP), CAS number: 85-86-7
Di-butyl phthalate (DBP), CAS number: 84-74-2

Sahin GERGIN
Sampling and Reporting
Date: 26.02.2021

Ertan ÖSTUNEL
PPE Laboratory Responsible
Date: 26.02.2021

Confirmed/
25.02.2021
Volkan AKIN
Laboratory Manager

TMMOB 25.02.2021 06:42:19

MNA Laboratory Manager: Yrd. Doç. Dr. Mustafa Kemal ÖZKAN
info@mnaag.com.tr | 0312 674 00 00 | Fax: 0312 674 22 11

**BELGE**


**AGL GRUP PETROL
KOZMETİK MEDİKAL İNŞ.GIDA SAN. TİC. LTD. ŞTİ**

**ELDİVEN
TEST RAPORU
EN 455
(GLOVE TEST REPORT)
EN 455**

RAPOR NO	SAYFA SAYISI	TEST /MUAYENE TARİHİ	RAPOR TARİHİ
21-G-PPE155222_R01	14	15.19.02.2021	25.02.2021

		Belge Mühendürlük Müayene - Teknik Kontrol Denetim ve Belgelendirme Hizmetleri		AGAB-T-013 21-G- PPE155222_R01 02/2021
		TS EN ISO/IEC 17025:2017		
		02/2021 Belge Mühendürlük - Müayene - Teknik Kontrol Denetim ve Belgelendirme Hizmetleri Ltd Şti		
		+90 850 888 0254 - +90 850 888 0816 info@belgesertifikasyon.com.tr		
TEST RAPORU - TESTING REPORT				
FİRMA BİLGİLERİ				
Müşteri Adı Client Name	AGL GRUP PETROL KOZMETİK MEDİKAL İNŞ.GIDA SAN. TİC. LTD. ŞTİ	Adres Address	Tugayyolu Caddesi, No:88, 6 Blok, Kat :15 Maltepe/İstanbul	
Firma Yetkilisi	Emre AĞAOĞLU	Teknik No	PPE155222	
Bağışım Contact	+90 533 013 11 13	e-mail	info@aglgrup.com.tr	
Telefon	+90 (216) 457 13 00	Test Tarihi	15-19.02.2021	
Fax	+90 (216) 457 13 00	Rapor No	21-G-PPE155222	
e-mail	info@aglgrup.com.tr	Rapor Tarihi	25.02.2021	
web	www.aglgrup.com.tr	Testi Yapılan Örnekler	Tek kullanımlık Pudrasız Eldiven (Disposable Powder-Free Glove)	
Marka	NITRAG	Model	S, L, M, XL (Mavi, Siyah) S, L, M, XL (Blue, Black)	
Rapor Tipi Report Type	<input type="checkbox"/> İlk First	<input type="checkbox"/> Ara Intermediate	<input checked="" type="checkbox"/> Final Final	
Revizyon Revision	00			
Açıklamalar Remarks				
Lot No/si	S202102-1512, M202102-15018, L202102-15017, XL202102-15016			
NITRAG Size:	S - M - L - XL (Mavi / Blue - Siyah/Black)			
AGL PETROL_R01	3			
<small>Bu Test raporunda verilen sonuçlar koruyucu eldivenler için geçerlidir. Test raporunu kullanmadan önce, koruyucu eldivenlerin kullanım talimatlarını okuyunuz. Bu rapor BELGE Mühendürlük - Müayene - Teknik Kontrol Denetim ve Belgelendirme Hizmetleri Ltd Şti tarafından hazırlanmıştır. İnceleme ve müdahale raporları geçerlidir. This report may not be reproduced and then used without the permission of the laboratory. Further reports and/or approvals will not be for valid. RAP-40 30.09.2020 Revizyon:04</small>				

 BELGE	Belge Mühendislik Malzeme – Teknik Kontrol Denetim ve Belgelendirme Hizmetleri TS EN ISO/IEC 17025:2017	AKA8-T-013 Z1-G- RPE155222_R01 02/2021
	TEST / KURULUŞU / TESTING COMPANY	
ADI, GÖREVİ / NAME AND POSITION HAZIRLAYAN / WRITTEN BY Önder KÜÇÜKALP Deney Personeli	TARİH / DATE 15.02.2021	İMZA / SIGNATURE 
MÜŞTERİ KURULUŞU / CUSTOMER COMPANY AGL GRUP PETROL KOZMETİK MEDİKAL İNŞ. GIDA SAN. TİC. LTD. ŞTİ		
ADI, GÖREVİ / NAME AND POSITION ONAYLAYAN / APPROVED BY Emre AĞAOĞLU	TARİH / DATE 15.02.2021	İMZA / SIGNATURE 
AGL PETROL_R01 <small> Bu Test raporuna verilen sonuçlar belirtilen testleri içeren parçılarda kontrol gerektiren ürünlerdir. Kontrol ve rapora ait sonuçlar, test sonucu olarak geçerlidir. Bu rapor Belge Mühendislik – Malzeme – Teknik Kontrol Denetim ve Belgelendirme Hizmetleri'nin yasal bir üstadıdır. Ürünler için belgenin uygulanması için lütfen müşteriye danışınız. İnceleme ve malzeme raporları için belgenin uygulanması için lütfen müşteriye danışınız. Testing report without signature and seal are not valid. RFP-49 10/06/2020 Reviz 04 </small>		

		Belge Mühendilik Müayene - Teknik Kontrol Denetim ve Belgelendirme Hizmetleri		ASAB T-011 21-0- PPE153222_R01 02/2021			
		TS EN ISO/IEC 17025:2017					
TEST TARİHİ / TEST DATE: 15-19.02.2021							
MÜŞTERİ KURULUŞ CLIENT COMPANY:							
AGL GRUP PETROL KOZMETİK MEDİKAL İNŞ.GIDA SAN. TİC. LTD. ŞTİ		Marka: Brand:	Modeli Model:	Seri No: Serial No:			
NITRAG		Tek kullanımlık Eldiven (Disposable Powder-Free Glove)		LOT No:			
İMALATÇISI MANUFACTURER:							
AGL GRUP PETROL KOZMETİK MEDİKAL İNŞ.GIDA SAN. TİC. LTD. ŞTİ		Marka: Brand:	Modeli Model:	Seri No: Serial No:			
NITRAG		Tek kullanımlık Eldiven (Disposable Powder-Free Glove)		LOT No:			
TEST ÖRNEKLERİ AÇIKLAMASI DESCRIPTION OF SAMPLES:							
No	Ürün Açıklaması Product Description	Marka Model Brand/ Model	Renk Colour	Parti No: Lot No	Büyük Size	Örnek Adedi Sample received (pieces)	Üretici Manufacturer
1	Tek kullanımlık Eldiven (Disposable Powder-Free Glove)	NITRAG	Mavi Blue	S202102-15012	S	300	AGL GRUP PETROL KOZMETİK MEDİKAL İNŞ.GIDA SAN. TİC. LTD. ŞTİ
				M202102-15018	M	300	
				L202102-15016	L	300	
				XL202102-15017	XL	300	
				Malzeme / Material: PE			
TEST ÖRNEKLERİ AÇIKLAMASI DESCRIPTION OF SAMPLES:							
No	Ürün Açıklaması Product Description	Marka Model Brand/ Model	Renk Colour	Parti No: Lot No	Büyük Size	Örnek Adedi Sample received (pieces)	Üretici Manufacturer
1	Tek kullanımlık Eldiven (Disposable Powder-Free Glove)	NITRAG	Siyah Black	S202102-15012	S	300	AGL GRUP PETROL KOZMETİK MEDİKAL İNŞ.GIDA SAN. TİC. LTD. ŞTİ
				M202102-15018	M	300	
				L202102-15016	L	300	
				XL202102-15017	XL	300	
				AGL PETROL_R01			
Bu Test raporundaki verilerin sonuçları belgeleri teknik olarak geçerli olarak göstermektedir. Kontrol ve testler için gerekli olan tüm ekipmanlar, ölçümler ve metodolojiler düzenli olarak kalibrasyon ve doğrulama için kullanılmaktadır. The report data can be re-evaluated after 30 days after the date of the test report. Testing results without signature and seal are not valid. RAP-02-10.08.2020 Revizyon:04							

	Belge Mühürsüz Muayene - Teknik Kontrol Denetim ve Belgelendirme Hizmetleri TS EN ISO/IEC 17025:2017		ALAB-001 21-6- PPE15322_R01 02/2021				
	Malzeme / Materyal PE						
Müşteri tarafından belirtilen lot büyüklüğü: Lot size as specified by client		15.001 ile 50.000 parça 15,001 to 50,000 pieces					
TEST YÖNTEMİ: METHOD OF TEST							
EN 455-1	Tek kullanımlık tıbbi eldivenler Bölüm 1, Deliklilik için gereksinimler ve test	Medical gloves for single use Part 1, Requirements and testing for freedom from holes					
EN 455-2	Tek kullanımlık tıbbi eldivenler Bölüm 2, Fiziksel özellikler için gereksinimler ve test	Medical gloves for single use Part 2, Requirements and testing for physical properties					
EN 455-3	Tek kullanımlık tıbbi eldivenler Bölüm 3, Biyolojik değerlendirme için gereksinimler ve test	Medical glove for single use Part 3, Requirements and testing for biological evaluation					
EN 455-4	Tek kullanımlık tıbbi eldivenler Bölüm 4, Raf ömrünün tayini için kanallar ve deneyler	Medical glove for single use Part 4, Rules and tests for determination of shelf life					
93/42 EC	Tıbbi Cihaz Direktifi	Medical Device Directive					
2016/425 EC	Kişisel Koruyucu Ekipman Direktifi	PPE Directive					
ASTM D6319 - 19	Tıbbi Uygulama için Nitril Muayene Eldivenleri için Standart	Standard Specification for Nitrile Examination Gloves for Medical					
ASTM D5351 - 19	Tıbbi Eldivenlerde Deliklerin Tespiti için Standart Test Yöntemi	Standard Test Method for Detection of Holes in Medical Gloves					
ASTM D6124 - 06	Tıbbi Eldivenler Üretimdeki Arızalar için Standart Test Yöntemi	Standard Test Method for Visual Powder on Medical Gloves					
SONUÇLAR RESULTS							
Tablo - 1 EN 455-1 Sonuçları Table - 1 Results for EN 455-1							
Madde Clause	Test Test	Boyut Size	Şartlar Requirements	Kabul Edilebilir Uyumuz Sayısı non-compliers allowed (pieces)	Test Edilen Numune Sayısı tested (pieces)	Karşılaştırma Değer Sayısı non-compliers found (pieces)	Sonuç Colored result
4. 5.	Serbest Delik Freedom from holes	S	Sızıntı yapmayacak Shelf not leak	1	50	1	Geçti Passed
		M		1	50	1	Geçti Passed
		L		1	50	1	Geçti Passed
		XL		1	50	1	Geçti Passed
AGL PETROL_R01				6			
Bu Test raporuna verilen sonuçlar belirtilen şartları yerine getirdiği kontrolü gerektirmediği sürece geçerlidir. Kontrol ve testlere ait sonuçlar test edilen ürünün kalitesini yansıtmaz. İzlenen ve ölçülen her süreç için geçerlidir. This report shall not be reproduced other than in full accord with the permission of the laboratory. Testing rooms without signature and seal are invalid. RAP-49-10.06.2020 Reviz:04							

7 days at (70±2)°C		XL	26	9,6	Geçti Passed																														
<p>SONUÇLAR RESULTS</p> <p>Tablo - 3 EN 455-2 Madde 7 Sonuçları Table-3 Results for EN 455-2 Clause 7</p> <table border="1"> <thead> <tr> <th>Maddes Clause</th> <th>Test Test</th> <th>Şartlar Requirements</th> <th>Sonuç Results</th> <th>Çıkarılan Sonuç Inferred results</th> </tr> </thead> <tbody> <tr> <td>7</td> <td>Etiketleme Labeling</td> <td>EN ISO 15223-1 ve EN 1041 uyarınca İmalatçılar eldiveni ve / veya orijin tarihi ile paketlenen Üretim tarihi paketine tarihi olarak tanımlar. Manufacturers shall label the glove and/or the packaging with the date of manufacture in accordance with EN ISO 15223-1 and EN 1041. Date of manufacture is defined as the packaging date.</td> <td>Gözlemlenen Observed</td> <td>Geçti Passed</td> </tr> </tbody> </table>						Maddes Clause	Test Test	Şartlar Requirements	Sonuç Results	Çıkarılan Sonuç Inferred results	7	Etiketleme Labeling	EN ISO 15223-1 ve EN 1041 uyarınca İmalatçılar eldiveni ve / veya orijin tarihi ile paketlenen Üretim tarihi paketine tarihi olarak tanımlar. Manufacturers shall label the glove and/or the packaging with the date of manufacture in accordance with EN ISO 15223-1 and EN 1041. Date of manufacture is defined as the packaging date.	Gözlemlenen Observed	Geçti Passed																				
Maddes Clause	Test Test	Şartlar Requirements	Sonuç Results	Çıkarılan Sonuç Inferred results																															
7	Etiketleme Labeling	EN ISO 15223-1 ve EN 1041 uyarınca İmalatçılar eldiveni ve / veya orijin tarihi ile paketlenen Üretim tarihi paketine tarihi olarak tanımlar. Manufacturers shall label the glove and/or the packaging with the date of manufacture in accordance with EN ISO 15223-1 and EN 1041. Date of manufacture is defined as the packaging date.	Gözlemlenen Observed	Geçti Passed																															
<p>SONUÇLAR RESULTS</p> <p>Tablo - 4 EN 455-3 Madde 4.2 - 4.5 Sonuçları Table-4 Results for EN 455-3 Clauses 4.2 - 4.5</p> <table border="1"> <thead> <tr> <th>Maddes Clause</th> <th>Test Test</th> <th>Şartlar Requirements</th> <th>Sonuç Açıklaması Results / Remarks</th> <th>Çıkarılan Sonuç Inferred results</th> </tr> </thead> <tbody> <tr> <td>4.2</td> <td>Kiryaşalları Chemicals</td> <td>Eldivende Talk Pudrası olmamasıdır. (Magnesyum silikati) Glove shall not be dusted with talcum powder. (magnesium silicate)</td> <td>Müşteri beyanına göre pudrasız eldivendir. Glove is powder free glove, based on client's declaration letter version</td> <td>Uygunmu N/A</td> </tr> <tr> <td></td> <td></td> <td>Diğer Kiryaşalları Other chemicals</td> <td>Kiryaşarıl Madde Listesi Müşteri tarafından bildirilecektir. Manufacturers shall declare upon request a list of chemical ingredients</td> <td>Uygunmu N/A</td> </tr> <tr> <td>4.3</td> <td>Endotoksinler Endotoxins</td> <td><28 EU/Çift Düşük endotoksin içeriği etiketlenmiş < 20 EU/par for gloves labeled with 'low endotoxin content'</td> <td>"Düşük" olarak etiketlenmemiş endotoksin içeriği Not labeled with 'low endotoxin content'</td> <td>Uygunmu N/A</td> </tr> <tr> <td>4.4 5.2</td> <td>Pudrasız Eldivenler Powder-free gloves</td> <td>Pudrasız Eldiven için Toplam toz kalitesi 2 mg geçmeyecektir. For powder-free gloves, the total quantity of powder residues shall not exceed 2 mg per glove.</td> <td>S 0.052 mg Eldiven başına/par glove M 0.051 mg Eldiven başına/par glove L 0.31 mg Eldiven başına/par glove XL 0.35 mg Eldiven başına/par glove</td> <td>Geçti Passed Geçti Passed Geçti Passed</td> </tr> <tr> <td>4.5 5.3</td> <td>Proteinler, Sızılabilir Protein, leakable</td> <td>İmalatçı doğal kauçuk lateks içeren eldivenler için sızılabilir protein seviyesini en aza indirmeli. The manufacturer shall strive to minimize the leachable protein level for gloves containing natural rubber latex.</td> <td>Doğal olmayan kauçuk lateks eldiven Non-natural rubber latex glove</td> <td>Uygunmu N/A</td> </tr> </tbody> </table>						Maddes Clause	Test Test	Şartlar Requirements	Sonuç Açıklaması Results / Remarks	Çıkarılan Sonuç Inferred results	4.2	Kiryaşalları Chemicals	Eldivende Talk Pudrası olmamasıdır. (Magnesyum silikati) Glove shall not be dusted with talcum powder. (magnesium silicate)	Müşteri beyanına göre pudrasız eldivendir. Glove is powder free glove, based on client's declaration letter version	Uygunmu N/A			Diğer Kiryaşalları Other chemicals	Kiryaşarıl Madde Listesi Müşteri tarafından bildirilecektir. Manufacturers shall declare upon request a list of chemical ingredients	Uygunmu N/A	4.3	Endotoksinler Endotoxins	<28 EU/Çift Düşük endotoksin içeriği etiketlenmiş < 20 EU/par for gloves labeled with 'low endotoxin content'	"Düşük" olarak etiketlenmemiş endotoksin içeriği Not labeled with 'low endotoxin content'	Uygunmu N/A	4.4 5.2	Pudrasız Eldivenler Powder-free gloves	Pudrasız Eldiven için Toplam toz kalitesi 2 mg geçmeyecektir. For powder-free gloves, the total quantity of powder residues shall not exceed 2 mg per glove.	S 0.052 mg Eldiven başına/par glove M 0.051 mg Eldiven başına/par glove L 0.31 mg Eldiven başına/par glove XL 0.35 mg Eldiven başına/par glove	Geçti Passed Geçti Passed Geçti Passed	4.5 5.3	Proteinler, Sızılabilir Protein, leakable	İmalatçı doğal kauçuk lateks içeren eldivenler için sızılabilir protein seviyesini en aza indirmeli. The manufacturer shall strive to minimize the leachable protein level for gloves containing natural rubber latex.	Doğal olmayan kauçuk lateks eldiven Non-natural rubber latex glove	Uygunmu N/A
Maddes Clause	Test Test	Şartlar Requirements	Sonuç Açıklaması Results / Remarks	Çıkarılan Sonuç Inferred results																															
4.2	Kiryaşalları Chemicals	Eldivende Talk Pudrası olmamasıdır. (Magnesyum silikati) Glove shall not be dusted with talcum powder. (magnesium silicate)	Müşteri beyanına göre pudrasız eldivendir. Glove is powder free glove, based on client's declaration letter version	Uygunmu N/A																															
		Diğer Kiryaşalları Other chemicals	Kiryaşarıl Madde Listesi Müşteri tarafından bildirilecektir. Manufacturers shall declare upon request a list of chemical ingredients	Uygunmu N/A																															
4.3	Endotoksinler Endotoxins	<28 EU/Çift Düşük endotoksin içeriği etiketlenmiş < 20 EU/par for gloves labeled with 'low endotoxin content'	"Düşük" olarak etiketlenmemiş endotoksin içeriği Not labeled with 'low endotoxin content'	Uygunmu N/A																															
4.4 5.2	Pudrasız Eldivenler Powder-free gloves	Pudrasız Eldiven için Toplam toz kalitesi 2 mg geçmeyecektir. For powder-free gloves, the total quantity of powder residues shall not exceed 2 mg per glove.	S 0.052 mg Eldiven başına/par glove M 0.051 mg Eldiven başına/par glove L 0.31 mg Eldiven başına/par glove XL 0.35 mg Eldiven başına/par glove	Geçti Passed Geçti Passed Geçti Passed																															
4.5 5.3	Proteinler, Sızılabilir Protein, leakable	İmalatçı doğal kauçuk lateks içeren eldivenler için sızılabilir protein seviyesini en aza indirmeli. The manufacturer shall strive to minimize the leachable protein level for gloves containing natural rubber latex.	Doğal olmayan kauçuk lateks eldiven Non-natural rubber latex glove	Uygunmu N/A																															
<p>AGL PETROL_R01</p> <p>Bu Test raporunda verilen sonuçlar belirlenen şartları karşılar ve şartlarında herhangi bir değişiklik yapılmamıştır. Ancak bu rapora ek olarak, test edilen ürünün paketleri için rapor Belge Mühürü - Beyan - Teknik Kontrol Denetim ve Belgelendirme Hizmetleri tarafından hazırlanmıştır. İnceleme raporları raporla birlikte sunulmaktadır. İnceleme raporları raporla birlikte sunulmaktadır.</p> <p>The report shall not be issued nor used than what is stated and the permission of the laboratory. Testing reports without copy of test and per set will not be issued. RAP-49 30-09-2020 Rev.04</p>																																			

		Belgę Mühendüklük Muayene – Teknik Kontrol Denetim ve Belgelendirme Hizmetleri TS EN ISO/IEC 17025:2017	ALAB-F011 33-G- PPE155222_R01 02/2021
SONUÇLAR RESULTS			
Tablo – 5 EN 455-3 Madde 4.6 Sonuçları Table-5 Results for EN 455-3 Clauses 4.6			
Madde Clause	Test Test	Kırtlar Assessments EN 1041 uyarınca Etiketlemeye ek olarak EN ISO 15223-1 uyarınca semboller aşağıdaki gereksinimleri karşılamalıdır. In addition to the labelling specified in EN 1041 and the relevant symbols given in EN ISO 15223-1, the following requirements apply: a) medical gloves containing natural rubber latex shall be labelled on the packaging of at least the smallest packaging unit with the EN ISO 15223-1 symbol for latex. The labelling shall include the following or equivalent warning statement together with the symbol: "Product contains natural rubber latex which may cause allergic reactions, including anaphylactic reactions". b) the labelling shall include a prominent indication of whether the glove is powdered or powder-free. c) sterile powdered gloves shall be labelled with the following or equivalent, CAUTION: Surface powder shall be removed aseptically prior to undertaking operative procedures, in order to minimize the risk of adverse tissue reactions. d) for any medical glove containing natural rubber latex the product labelling shall not include: - any term suggesting relative safety, such as low allergenicity, hypoallergenicity or low protein. - any unjustified indication of the presence of allergens. e) if the manufacturer labels the gloves with the protein content, the protein limit, measured as specified in 5.3 shall be given.	Sonuç Results Uygulanmaz N/A Uygulanmaz N/A Uygun Comply Uygulanmaz N/A Uygulanmaz N/A Uygulanmaz N/A
4.6	Etiketleme Labeling		
		Çıkarılan Sonuç Annotated results	Geçti Passed
Testi Yapan Deney Personeli İsim – Soyisim / İmza Önder KÜÇÜKALP 25.02.2021		Raporu Hazırlayan İsim – Soyisim / İmza Önder KÜÇÜKALP 25.02.2021	Onaylayan Teknik Müdür İsim – Soyisim / İmza Aycaşu ÇELİKUSOY 25.02.2021
AGL PETROL RDJ Bu Test raporunda verilen sonuçlar belgeden herhangi bir şekilde alınmış ve değiştirilmeden raporlanmıştır. Kontrol ve onayları aynı şekilde yapılmalıdır. Bu rapor Belgę Mühendüklük – Muayene – Teknik Kontrol Denetim ve Belgelendirme Hizmetleri ile aynı form düzeninde kontrol edilmiştir. Kopyaların yapılması yasaktır. This report shall not be reproduced from this CD except with the permission of the client. Testing reports without signature and date are not valid. RAP-29-10.06.2020 Rev.04			

 BELGE	Belge Mühendislik Muayene – Teknik Kontrol Denetim ve Belgelendirme Hizmetleri	ALAB T-011 23-0- PPE155222_R01 02/2021
	TS EN ISO/IEC 17025:2017	

Deney ve /veya ölçüm sonuçları, gerçekleştirilmiş ölçüm belgeleri (olması halinde) ve deney metotları bu sertifikasyon talebinde belirtilen tüm teknik şartnamelerde yer almaktadır. The test and/or measurement results, the uncertainties (if applicable) with confidence probability and test methods are given on the following pages which are part of this report

Açıklamalar:
Remarks:

- S, M, L, XL, Deliksizlik testi müşteri beyanına göre yapılmıştır.
Freedom from holes test for S, M, L, XL, were tested in manufacturer's declaration
- Etiketleme gereksinimleri müşteri beyanı ile sunulan ambalaj resimine göre değerlendirilmiştir.
Labeling requirements are assessed based on submitted packaging artwork together with client's declaration
- Uygulanamaz: Gönderilen numune için geçerli değildir.
NA: Not applicable for the submitted sample.

Önder KÜÇÜKALP  Deney Personeli/Rapor Sorumlusu	Ayşegül CANLIŞOY  Teknik Müdür
---	---






Tek kullanımlık Pedrasız Eldiven Mavi - Black (NITRAG - S202102 - 15018)
 (Disposable Powder Free Glove) Mavi - Black (NITRAG - S202102 - 15018)

AGL PETROL_R01 10

Doğru Test Sonuçları İçin Lütfen Testin Gerçekleştirilmesini Sağlayan Kurumunuzu Bildiriniz. Doğru Test Sonuçları İçin Lütfen Testin Gerçekleştirilmesini Sağlayan Kurumunuzu Bildiriniz. For Correct Test Results, Please Inform the Institution that Ensures the Correct Test Results. For Correct Test Results, Please Inform the Institution that Ensures the Correct Test Results.


BU BELGE, BELGE MÜHÜRÜ VE/VEYA BELGE NOYLA SAĞLAMLIK İZLENİMİ YARATMAK İÇİN KULLANILMAMALIDIR. BU BELGE, BELGE MÜHÜRÜ VE/VEYA BELGE NOYLA SAĞLAMLIK İZLENİMİ YARATMAK İÇİN KULLANILMAMALIDIR. THIS DOCUMENT SHOULD NOT BE USED TO CREATE AN IMPRESSION OF RELIABILITY. THIS DOCUMENT SHOULD NOT BE USED TO CREATE AN IMPRESSION OF RELIABILITY.

RAP.49 10.06.2020 Rev.04

 BELGE	Belge Mühendisi Muayene – Teknik Kontrol Denetim ve Belgelendirme Hizmetleri	AGL-T-011 22-G PPE155222_R01 02/2021
TS EN ISO/IEC 17025:2017		

*** Rapor Sonu / End of Report ***

AGL GRUP


 **BELGE**
MUAYENE

BÖLÜM 2 / SECTION 2
PROTOKOL/PROTOCOL

AGL GRUP PETROL
KOZMETİK MEDİKAL İNŞ.GIDA SAN. TİC. LTD. ŞTİ

ELDİVEN
TEST RAPORU
EN 455
(GLOVE TEST REPORT)
EN 455

AGL PETROL_R01 11



Bu Test raporunda verilen sonuçlar laboratuvar koşullarında, müşteri tarafından kontrol edilmiş ve doğrulanmış ürünlerdir. Müşteri, laboratuvar ve testler için gerekli tüm koşulları sağlar. Bu rapor Belge Mühendisi – Muayene – Teknik Kontrol Denetim ve Belgelendirme Hizmetleri tarafından hazırlanmıştır. Müşteri ve müşteri raporunu gözetir.

This report shall not be reproduced or further used without the permission of the laboratory. Testing room safety regulations apply and are valid. RAP.49 10.06.2020 Rev.04

	Belgə Mühəndislik Müayinə – Teknik Kontrol Dənətin və Səpələrdənə Hizmetleri	ALAB 1101 31-G PPE15522_001 02/2021
	TS EN ISO/IEC 17025:2017	
	<p>ekvivalenti 70 ° C'de 3 gün (veya) veya 7 gün (Nitril, lateks) vaxtdır. 15-96 saat dənətməyə buraxılıb, sonra təst edilir. Jensle test, elastomeric materials (ASTM D412) / Glove tensile test, aged and unaged (ASTM D3576). The test specimen is cut with the ASTM D412 die. Çətinlik testi, elastomeric maddələr (ASTM D412) / Glove tensile test, aged and unaged (ASTM D3576). The test specimen is cut with the ASTM D412 die. Latex, lateks və digər elastiya maddələr üçün 15-96 saat dənətməyə buraxılıb, sonra təst edilir. Jensle test, elastomeric materials (ASTM D412) / Glove tensile test, aged and unaged (ASTM D3576). The test specimen is cut with the ASTM D412 die.</p> <p>Tükənlü vəirətli xüsusiyyətlər (NPA 1996) Bu test prosedürü, bütün əlverişli və əl barları istifadəsi üçün dəyərli məlumatlar təqdim edir. İstehlakçıların 100 bar sürətləndirilmiş çətinlik (PHE334) testindəki bir dəfəlik yığılma və kəmərlə kəmərləndirilmə 100 bar sürətli çətinlik testinə bənzərdir. Təst nümunəsi, dənədə 100 - 115 dəvə (saniyə) ilə çətinliklərdən ibarət çətinliklərdə 1 saatlıq bir təhlükəli olma qüvvəsi ilə müdafiə və daha sonra xüsusi xüsusiyyətlər üçün təhlükəlidir. Whistle glove viral barrier (NPA 1996). This test procedure is used for evaluating viral barrier properties of whole gloves. Gloves are suspended in a viral fluid test subject in the challenge vessel (PHE334) hastenəndirilmiş at a concentration of 210⁷ plaque forming units (PFU) per ml. The test sample are exposed throughout a 1-hour challenge on an orbital shaker operated at 100 - 115 revolutions per minute (RPM) and then assayed for viral penetration.</p> <p>Dənətmə vəirətli xüsusiyyətləri (ASTM D571, NPA 1996) ASTM D571-06, test nümunəsi bir-iyirmi 100 x 2 saat 20 ± 3 ° C'də; Kəmərləndirilmə, NPA 1996 (süra, test nümunəsi bir-iyirmi 100 x 2 ° C'də 22 ± 0.3 saat) vaxtdır. Glove heat aging degradation test (ASTM D571, NPA 1996). For ASTM D571, the test specimen is conditioned in an oven at 70 ± 2°C for 168 ± 2 hours. For NPA 1996, the test specimen is conditioned in the oven at 100 ± 2°C for 22 ± 0.3 hours.</p> <p>İstehlakçılara məlumatlar (ASTM D551) Bu test, standart və lazımı müayinə və / vəya cəmiyyət əhəmiyyətli istifadə üçün bütün və eyni test prosedürü ilə edilmişdir. Whistle glove viral barrier (NPA 1996) Bu test prosedürü, bütün əlverişli və əl barları istifadəsi üçün dəyərli məlumatlar təqdim edir. İstehlakçıların 100 bar sürətləndirilmiş çətinlik (PHE334) testindəki bir dəfəlik yığılma və kəmərlə kəmərləndirilmə 100 bar sürətli çətinlik testinə bənzərdir. Təst nümunəsi, dənədə 100 - 115 dəvə (saniyə) ilə çətinliklərdən ibarət çətinliklərdə 1 saatlıq bir təhlükəli olma qüvvəsi ilə müdafiə və daha sonra xüsusi xüsusiyyətlər üçün təhlükəlidir. Whistle glove viral barrier (NPA 1996). This test procedure is used for evaluating viral barrier properties of whole gloves. Gloves are suspended in a viral fluid test subject in the challenge vessel (PHE334) hastenəndirilmiş at a concentration of 210⁷ plaque forming units (PFU) per ml. The test sample are exposed throughout a 1-hour challenge on an orbital shaker operated at 100 - 115 revolutions per minute (RPM) and then assayed for viral penetration.</p> <p>Bədənə dənətmə (ASTM F1342) Bu test, yüngül, kəskin kənarlı bir cəmiyyət problemlərini müəyinə etmək üçün istifadə edilən bir testdir. Whistle glove viral barrier (NPA 1996) Bu test prosedürü, bütün əlverişli və əl barları istifadəsi üçün dəyərli məlumatlar təqdim edir. İstehlakçıların 100 bar sürətləndirilmiş çətinlik (PHE334) testindəki bir dəfəlik yığılma və kəmərlə kəmərləndirilmə 100 bar sürətli çətinlik testinə bənzərdir. Təst nümunəsi, dənədə 100 - 115 dəvə (saniyə) ilə çətinliklərdən ibarət çətinliklərdə 1 saatlıq bir təhlükəli olma qüvvəsi ilə müdafiə və daha sonra xüsusi xüsusiyyətlər üçün təhlükəlidir. Whistle glove viral barrier (NPA 1996). This test procedure is used for evaluating viral barrier properties of whole gloves. Gloves are suspended in a viral fluid test subject in the challenge vessel (PHE334) hastenəndirilmiş at a concentration of 210⁷ plaque forming units (PFU) per ml. The test sample are exposed throughout a 1-hour challenge on an orbital shaker operated at 100 - 115 revolutions per minute (RPM) and then assayed for viral penetration.</p> <p>Əyri və yağlıq test (ASTM D6124) Bu test, əlverişli vəirətli çətinlik testini aşdıraraq, test nümunəsinin qırıqlıq qüvvəsinə bənzərdir. Whistle glove viral barrier (NPA 1996) Bu test prosedürü, bütün əlverişli və əl barları istifadəsi üçün dəyərli məlumatlar təqdim edir. İstehlakçıların 100 bar sürətləndirilmiş çətinlik (PHE334) testindəki bir dəfəlik yığılma və kəmərlə kəmərləndirilmə 100 bar sürətli çətinlik testinə bənzərdir. Təst nümunəsi, dənədə 100 - 115 dəvə (saniyə) ilə çətinliklərdən ibarət çətinliklərdə 1 saatlıq bir təhlükəli olma qüvvəsi ilə müdafiə və daha sonra xüsusi xüsusiyyətlər üçün təhlükəlidir. Whistle glove viral barrier (NPA 1996). This test procedure is used for evaluating viral barrier properties of whole gloves. Gloves are suspended in a viral fluid test subject in the challenge vessel (PHE334) hastenəndirilmiş at a concentration of 210⁷ plaque forming units (PFU) per ml. The test sample are exposed throughout a 1-hour challenge on an orbital shaker operated at 100 - 115 revolutions per minute (RPM) and then assayed for viral penetration.</p> <p>Fiziki ölçmə testləri (ASTM D3576-01, ASTM D375, ASTM D3556) Bu test əlverişli vəirətli ölçmə testlərini müəyinə edir. Whistle glove viral barrier (NPA 1996) Bu test prosedürü, bütün əlverişli və əl barları istifadəsi üçün dəyərli məlumatlar təqdim edir. İstehlakçıların 100 bar sürətləndirilmiş çətinlik (PHE334) testindəki bir dəfəlik yığılma və kəmərlə kəmərləndirilmə 100 bar sürətli çətinlik testinə bənzərdir. Təst nümunəsi, dənədə 100 - 115 dəvə (saniyə) ilə çətinliklərdən ibarət çətinliklərdə 1 saatlıq bir təhlükəli olma qüvvəsi ilə müdafiə və daha sonra xüsusi xüsusiyyətlər üçün təhlükəlidir. Whistle glove viral barrier (NPA 1996). This test procedure is used for evaluating viral barrier properties of whole gloves. Gloves are suspended in a viral fluid test subject in the challenge vessel (PHE334) hastenəndirilmiş at a concentration of 210⁷ plaque forming units (PFU) per ml. The test sample are exposed throughout a 1-hour challenge on an orbital shaker operated at 100 - 115 revolutions per minute (RPM) and then assayed for viral penetration.</p>	
AGL PETROL_001 13		

The registered test is repeated after the test is completed with the permission of the customer. Test reports without signature and seal are not valid. RAP-49 10.06.2020 Reviziyası

	Belge Mühendislik Muayene – Teknik Kontrol Denetim ve Belgelendirme Hizmetleri		ALAB 5-001 25-G PPE150222_R01 02/2021
	TS EN ISO/IEC 17025:2017		

Table 2 – Median values of force at break

	Force at break in N/100		
	Surgical gloves a)	Examination/procedure gloves b) c)	
Throughout shelf life tested according to 5.2 and within 12 months of manufacture tested according to 5.3	≥ 8.0	≥ 6.0	≥ 3.0

a) Requirements for all surgical gloves.
 b) Requirements for all examination gloves, except gloves made from transparent material (e.g. polyurethane) or polyethylene.
 c) Requirements for gloves made from transparent material (e.g. polyurethane) or polyethylene.

Please note that this Report is issued under the following terms :

- This report applies to the sample of the specific product/equipment given at the time of its testing/calibration. The results are not used to indicate or imply that they are applicable to other similar items. In addition, such results must not be used to indicate or imply that BELGE SERTİFİKASYON approves, recommends or endorses the manufacturer, supplier or user of such product/equipment, or that BELGE SERTİFİKASYON in any way "guarantees" the later performance of the product. Unless otherwise stated in this report, no tests were conducted to determine long term effects of using the specific product/equipment.
- The sample/s mentioned in this report /s/ are submitted/supplied/manufactured by the Client. BELGE SERTİFİKASYON therefore assumes no responsibility for the accuracy of information on the brand name, model number, origin of manufacture, consignment or any information supplied.
- Nothing in this report shall be interpreted to mean that BELGE SERTİFİKASYON has verified or ascertained any endorsement or marks from any other testing authority or bodies that may be found on that sample.
- This report shall not be reproduced wholly or in parts and no reference shall be made by the Client to BELGE SERTİFİKASYON or to the report or results furnished by BELGE SERTİFİKASYON in any advertisements or sales promotion.

KABUL KRİTERLERİNDE YAPILAN DEĞİŞİKLİKLER ve ONAY

YAPILAN DEĞİŞİKLİK DESCRIPTION of CHANGE	ÖNCEDEN BELİRLENMİŞ OLAN KABUL KRİTERİ PRE-DETERMINED ACCEPTANCE CRITERIA	YENİ KABUL KRİTERİ NEW ACCEPTANCE CRITERIA	DEĞİŞİKLİĞİN GEREKÇESİ REASON OF THE CHANGE

DEĞİŞİKLİKLERİN ONAYLANMASI / APPROVAL of THE CHANGES Yapılan değişiklikler protokola ek imzalayan kişiler tarafından onaylanacaktır / The changes must be approved by persons in charge of the laboratory.

AGL PETROL R01 14

AGL PETROL R01 14
 Bu Zimmetli raporun her bir kopyasının dağıtımını sağlamak amacıyla belgenin elektronik ortamına aktarılmasıdır. Kontrol ve teslimat için raporun her bir kopyasının dağıtımını sağlamak amacıyla belgenin elektronik ortamına aktarılmasıdır. İMZA ve MÜHÜRLEME GEREKMEZ.
 This report shall not be transmitted other than in full except with the permission of the laboratory. To avoid reports without signatures and stamps, the report is transmitted in electronic form. İMZA VE MÜHÜRLEME GEREKMEZ.
 RAP-40 10.06.2020 Rev.04

 BELGE	Belge Mühendislik Muayene – Teknik Kontrol Denetim ve Belgelendirme Hizmetleri	ALAB-1-001 21-G PPE155222_R01 02/2021
	TS EN ISO/IEC 17025:2017	
ONAYLAYAN ADI, GÖREVİ / APPROVAL NAME AND POSITION	TARİH / DATE	İMZA / SIGNATURE
TEST KURULUŞU TESTING COMPANY	Ayçaşu CANLIŞOY	25.02.2021
MÜŞTERİ KURULUŞ CLIENT COMPANY	Emre AĞAOĞLU	25.02.2021
		
		
		
		
		

AGL PETROL_R01

15

Bu Test raporunda verilen sonuçlar belirlenen yöntemlere göre belirlenmiş kontrol yöntemleri kullanılarak elde edilmiştir. Kontrol ve testlere ait süreler, test sonuçları ile birlikte bu rapor Belge Mühendislik - Muayene - Teknik Kontrol Denetim ve Belgelendirme Hizmetleri'ne veya bir çalışanımıza veya herhangi bir kişiye verilmemelidir. Başkası ve farklı amaçlar için raporlanmamalıdır.

This report shall not be reproduced or used for any other purpose without the permission of the laboratory. Testing results without signature and seal are not valid. RAP-09 10.06.2020 Rev 04



**EC DECLARATION OF CONFORMITY
AB UYGUNLUK BEYANI**

Üretici / Manufacturer
AGL GLOVE FİTİKOL KÜVEMİŞ MÜHÜRÜN
İNG. İZBA SAN. TİC. LTD. ŞTİ

Adres / Address
TUSAY YOLU CO. OFİSİN İSTANBUL PLAZA B BLOK NO: B8 KAT: 15
34341 MALTEPE, İSTANBUL, TÜRKİYE

Ürünün Markası / Brand of Product



Ürünün İsmi / Name of Product
ÇOK AMAÇLI TEK KULLANIMLI THE EL DÜVEN
MULTIPURPOSE DISPOSABLE THE GLOVES

Tip - Model / Type-Model
SINIF 1 - TİP 1 / CLASS 1 - TYPE 1

Harmonize Standartlar / Harmonized Standards

EN 420+A1:2010, EN ISO 379-1:2015
EN ISO 374-2:2013, EN ISO 374-3:2016, TS EN 16623-3+A1
TS EN 1394-4:2019, EN ISO 1394-5:2016, EN 455-1:2010,
EN 455-2:2015, EN 455-3:2015

Yönetmelik ve Yönetmelikler / Directives and Regulations

2007/47/EC KONDİSYON GEREKİMLİ - 2007/47/EC COUNCIL DIRECTIVE
93/42/EEC TIBBİ CİHAZLAR YÖNETMELİĞİ - 93/42/EEC MEDICAL DEVICES DIRECTIVE

The Certificate is issued under the following conditions:

1. It applies only to the above referenced models of the medical devices.
2. It does not imply that UKS has performed any surveillance or control of their manufacture.
3. The manufacturer is obligated to ensure conformity of all in-medical devices of the respective model to type assessment by the issue of this certificate.
4. The certificate remains valid until the manufacturing conditions, the quality system or relevant legislation are changed.
5. After building of the relevant EC legislation requirements, the manufacturer shall affix to each medical device, of the above referenced models, the CE marking according to 93/42 example.

Sertifika No / Certificate No: CE1-0506
Sertifika Tarihi / Certificate Date: 30.01.2022
Sertifika Baki / Expiration Date: 10.01.2023

Genel Müdür / General Manager



www.uksbelgelerdirme.com.tr

Bilgi için: iletisim@uksbelgelerdirme.com.tr adresine e-posta gönderilebilir.

UKS Uluslararası Kalite Sistemleri Ve Belgelendirme Ltd. Şti.
Anadoluhisari, Tuzla Cad. No:21, Beşiktaş / ZKMBİS
İstanbul - Türkiye
Telefon: +90 212 331 41 77 / Faks: +90 212 331 41 47
www.uksbelgelerdirme.com.tr



This is to certify

**Agl Grup Petrol Kozmetik Medikal
Ins. Gıda San. Tic. Ltd. Sti.**

**TUGAY YOLU CD. OFİSİM İSTANBUL PLAZA B BLOK NO.: 88 KAT: 16
34844 MALTEPE, İSTANBUL
TURKEY**

*Has been assessed by UKS and found to be in compliance
with the following standard*

ISO 10002:2018

*The Quality Management - Customer Satisfaction System
is applicable to:*

**Ffp2 Ffp3 Surgical Mask, Overalls, Apron, Nitrag TPE Examination Gloves, Laminated
Protective Boots Production**

Category / Sub Category	1 -
Certificate No	MSC-0026
Certificate First Issue Date	09 th February 2021
Decision Date	09 th February 2021
Certificate Issue Date	09 th February 2021
Expiration Date	08 th February 2022

This certificate is valid for 1 year after the submission of surveillance inspection data.
It is valid as long as it complies with ISO 9001:2015. The product status and its compliance with the certificate conditions is valid only in year.
UKS Reference No: 01.0001/0001.0001/0001.0001/0001.0001

General Manager



www.uksgelendime.com.tr

Please check the validity of the certificates from our web site

UKS International Quality Systems & Certification Ltd.
Mikser Sok. Dışarı Sok. No:21 Eminönü (İSTANBUL, TÜRKİYE)
Phone: +90 216 239 45 77 Fax: +90 216 239 47 47
www.uksgelendime.com.tr



This is to certify

**Agl Grup Petrol Kozmetik Medikal
Ins. Gıda San. Tic. Ltd. Sti.**

**TUGAY YOLU CD. OFİSİM İSTANBUL PLAZA B BLOK NO.: 88 KAT: 15
34844 MALTEPE, İSTANBUL
TURKEY**

*Has been assessed by UKS and found to be in compliance
with the following standard*

ISO 13485:2016

The Medical Devices Management System is applicable to;

Ffp2 Ffp3 Surgical Mask, Overalls, Apron, NitrAG TPE Examination Gloves, Laminated
Protective Boots Production

Category / Sub Category : ---
Certificate No : MDO-0116
Certificate First Issue Date : 09 February 2021
Decision Date : 09 February 2021
Certificate Issue Date : 09 February 2021
Expiration Date : 08 February 2022

Information reported is based on the information provided by the client and is not a guarantee of the accuracy of the information. UKS International Quality Systems & Certification Ltd. is not responsible for the accuracy of the information provided by the client.



General Manager



www.aksbelgedime.com.tr
Please check the validity of the certificate from our web site.

UKS International Quality Systems & Certification Ltd.
Akıncı Sok. 34. Kat, No:35 Beşiktaş / İSTANBUL / TÜRKİYE
Phone: +90 212 228 45 77 Fax: +90 212 228 47 47
www.aksbelgedime.com.tr



This is to certify

**AgI Grup Petrol Kozmetik Medikal
İns. Gıda San. Tic. Ltd. Sti.**

**TUGAY YOLU CD. OFİSİM İSTANBUL PLAZA B BLOK NO.: 88 KAT: 15
34844 MALTEPE, İSTANBUL
TURKEY**

*Has been assessed by UKS and found to be in compliance
with the following standard*

TS EN ISO 9001:2015

The Quality Management System is applicable to:

**Ffp2 Ffp3 Surgical Mask, Overalls, Apron, NitrAg TPE Examination Gloves, Laminated
Protective Boots Production**

Category / Sub-Category : —
Certificate No : MSQ-4265
Certificate First Issue Date : 09th February 2021
Decision Date : 09th February 2021
Certificate Issue Date : 09th February 2021
Expiration Date : 08th February 2022

This certificate is valid for 1 year after the verification or re-verification decision date.
In case of change of scope, UKS International Quality Systems & Certification Ltd. shall be notified in writing within 15 days.
UKS Process No: 01-01 (Revizyon) Tarihi: 14.03.2019 Etiler, İstanbul



General Manager



www.aksbelgelendirme.com.tr

Please check the validity of the certificate from our web site

UKS International Quality Systems & Certification Ltd.
Aksepet Park, Dışca Cad. No:21 Ümraniye - 05409001, TÜRKİYE
Türkiye : +90 216 228 41 77 / İngiltere : +44 216 228 41 77
info@uksbelgelendirme.com.tr



This is to certify

**AgI Grup Petrol Kozmetik Medikal
Ins. Gıda San. Tic. Ltd. Sti.**

**TUGAY YOLU CD. OFİSİM İSTANBUL PLAZA B BLOK NO.: 88 KAT: 15
34844 MALTEPE, İSTANBUL
TURKEY**

*Has been assessed by UKS and found to be in compliance
with the following standard*

**GMP
Good Manufacturing Practice**

The Good Manufacturing Practice System is applicable to:

**Ffp2 Ffp3 Surgical Mask, Overall, Apron, NitrAg TPE Examination Gloves, Laminated
Protective Boots Production**

Category / Sub Category : —
Certificate No. : MSU-0017
Certificate First Issue Date : 09th February 2021
Decision Date : 09th February 2021
Certificate Issue Date : 09th February 2021
Expiration Date : 08th February 2022

This certificate is valid for 3 years after the certification date unless the
scope of the certificate is extended. Please refer to the certificate for the
conditions of use. For more information, please contact us.

General Manager



www.uksbelgedirme.com.tr
Please check the validity of the certificate from our web site.

UKS International Quality Systems & Certification Ltd.
Anadoluhisari, SSK Blok. No:35 Kat:15 (İSTANBUL, TÜRKİYE)
Phone: +90 212 228 42 77 / Fax: +90 212 228 42 47
info@uksbelgedirme.com.tr



This is to certify

**AgI Grup Petrol Kozmetik Medikal
İns. Gıda San. Tic. Ltd. Şti.**

**TUGAY YOLU CD. OFİSİM İSTANBUL PLAZA B BLOK NO.: 88 KAT: 15
34844 MALTEPE, İSTANBUL
TURKEY**

*Has been assessed by UKS and found to be in compliance
with the following standard*

TS ISO 45001:2018

*The Occupational Health And Safety Management System
is applicable to:*

**Ffp2 Ffp3 Surgical Mask, Overalls, Apron, Nitrax TPE Examination Gloves, Laminated
Protective Boots Production**

Category / Sub Category : --
Certificate No. :OHS-0047
Certificate First Issue Date :09th February 2021
Decision Date :09th February 2021
Certificate Issue Date :09th February 2021
Expiration Date :08th February 2022

This certificate is valid for 1 year after the certification or re-certification date unless
in case of absence of records with ISO Certification Bodies it is extended to the
next issue certificate of valid date + one
UKS, Turkey Branch No: 0212 448 00 00, Sarıyer, İstanbul, Turkey



General Manager



www.ukselendime.com.tr
Please check the validity of the certificate from our web site

UKS International Quality Systems & Certification Ltd.
Akmer Sok. Sıkla Cad. No:35 Üsküdar / İSTANBUL / TÜRKİYE
Phone: +90 212 328 43 77 Fax: +90 212 328 47 47
www.ukselendime.com.tr



This is to certify

**AgI Grup Petrol Kozmetik Medikal
Ins. Gıda San. Tic. Ltd. Sti.**

**TUGAY YOLU CD. OFİSİM İSTANBUL PLAZA B BLOK NO.: 88 KAT: 15
34844 MALTEPE, İSTANBUL
TURKEY**

*Has been assessed by UKS and found to be in compliance
with the following standard*

TS EN ISO 22000:2018

The Food Safety Management System is applicable to:

Ffp2 Ffp3 Mask, Food Contact Overalls, Aprons, TPE Gloves Production

Category / Sub Category : JU-1
Certificate No : MSF-0353
Certificate First Issue Date : 12th February 2021
Decision Date : 12th February 2021
Certificate Issue Date : 12th February 2021
Expiration Date : 11th February 2022

This certificate is valid for 1 year after the verification or re-verification services are completed.
Scope of Accreditation: MSF-0353: Certification Body not accredited in the mentioned certificate in their scope of accreditation.
MSF-0353: Date of Issue: 12/02/2021 Date of Expiry: 11/02/2022



General Manager



www.ulsbelgelendirme.com.tr
Please check the validity of the certificate from our web site.

UKS International Quality Systems & Certification Ltd.
Andon Mah. Etiler Cad. No:33 Beşiktaş / İSTANBUL / TÜRKİYE
Phone: +90 212 339 43 77 Fax: +90 212 331 87 87
info@uksbelgelendirme.com.tr



This is to certify

**Agl Grup Petrol Kozmetik Medikal
Ins. Gıda San. Tic. Ltd. Sti.**

**TUGAY YOLU CD. OFİSİM İSTANBUL PLAZA B BLOK NO.: 88 KAT: 15
34844 MALTEPE, İSTANBUL
TURKEY**

*Has been assessed by UKS and found to be in compliance
with the following standard*

**GHP
Good Hygiene Practice**

The Good Hygiene Practice System is applicable to:

**Ffp2 Ffp3 Surgical Mask, Overalls, Apron, Nitril TPE Examination Gloves, Laminated
Protective Boots Production**

Category / Sub-Category : ---
Certificate No. : MSH-0005
Certificate First Issue Date : 09th February 2021
Decision Date : 09th February 2021
Certificate Issue Date : 09th February 2021
Expiration Date : 08th February 2022

This certificate shall be valid as long as the licensee complies with the requirements of the standard.

General Manager



www.uksbegelemdirme.com.tr

Please check the validity of the certificate from our web site.

UKS International Quality Systems & Certification Ltd.
Atatürk Sok. 35/4 Kat. No:35 Üsküdar / İSTANBUL / TÜRKİYE
Phone: +90 212 328 15 77 Fax: +90 212 331 67 47
info@uksbegelemdirme.com.tr



This is to certify

**AgI Grup Petrol Kozmetik Medikal
Ins. Gıda San. Tic. Ltd. Sti.**

**TUGAY YOLU CD. OFİSİM İSTANBUL PLAZA B BLOK NO.: 88 KAT: 15
34844 MALTEPE, İSTANBUL
TURKEY**

*Has been assessed by UKS and found to be in compliance
with the following standard*

TS EN ISO 14001:2015

The Environmental Management System is applicable to:

**Ffp2 Ffp3 Surgical Mask, Overalls, Apron, NitrAG TPE Examination Gloves, Laminated
Protective Boots Production**

Category / Sub Category : —
Certificate No. : MSE-0115
Certificate First Issue Date : 09th February 2021
Decision Date : 09th February 2021
Certificate Issue Date : 09th February 2021
Expiration Date : 08th February 2022

No signature is valid if a post date (signature) or a certification number date.
Dates of first or expiry of ISO 9001 Certification shall be determined by the certification body and shall be a part
of the certificate. For more information, see the website, www.uks.com



General Manager



www.ukselendirmasi.com.tr

Please check the validity of the certificate from our web site

UKS International Quality Systems & Certification Ltd.
Mühür Sok. 3/A Kat: No:33 Beşiktaş / İSTANBUL / TÜRKİYE
Phone: +90 212 332 43 77 Fax: +90 212 332 43 47
info@uksinternational.com.tr

AGL ^{GRUP}

at your service globally.

NitrAG[®]

Thank you for your time.



AGL GRUP Petrol Kozmetik Medikal İnş. Gıda San. Tic. Ltd. Şti.
Tugay Yolu Cd. B Blok No.: 88 Kat: 15, 34844 Maltepe, İstanbul, TURKEY
T +90 216 457 13 00 | info@aglgrup.com.tr

www.nitrage.com.tr

www.aglgrup.com.tr

SOFT 
TOUCH